

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2643

Client ID: SSTDCCC0.4

Date Analyzed: 07/22/2025

Lab File ID: BN037532.D

Time Analyzed: 10:51

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	7043	17.086	6014	21.277	5778	23.513
UPPER LIMIT	14086	17.586	12028	21.777	11556	24.013
LOWER LIMIT	3521.5	16.586	3007	20.777	2889	23.013
EPA SAMPLE NO.						
01 PB168952BL	6964	17.10	4782	21.28	4734	23.51
02 PB168952BS	4507	17.09	3057	21.27	2587 *	23.51
03 PB168952BSD	4034	17.09	3054	21.27	2644 *	23.51
04 RW7-SP100-20250717	5819	17.09	4394	21.28	3754	23.51
05 RW7-SP201-20250717	4844	17.09	3799	21.28	3347	23.51
06 RW7-SP302-20250717	4929	17.09	3886	21.28	3365	23.51
07 RW7-SP303-20250717	4425	17.09	3605	21.28	3019	23.51

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.