

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037531.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2643
DFTPP Injection Date: 07/22/2025
DFTPP Injection Time: 10:07

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	80.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.8 (21.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037532.D	07/22/2025	10:51
PB168952BL	PB168952BL	BN037533.D	07/22/2025	11:28
RW7-SP100-20250717	Q2643-01	BN037537.D	07/22/2025	13:52
RW7-SP201-20250717	Q2643-02	BN037538.D	07/22/2025	14:28
RW7-SP302-20250717	Q2643-03	BN037539.D	07/22/2025	15:04
RW7-SP303-20250717	Q2643-04	BN037540.D	07/22/2025	15:40
PB168952BS	PB168952BS	BN037544.D	07/22/2025	18:05
PB168952BSD	PB168952BSD	BN037545.D	07/22/2025	18:41
SSTDCCC0.4EC	SSTDCCC0.4	BN037546.D	07/22/2025	19:17