

LAB CHRONICLE

OrderID:	Q2644	OrderDate:	7/18/2025 11:21:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	O33

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2644-01	RW8-SP100-2025071 7	WATER			07/17/25 13:05			07/18/25
			TDS	SM2540 C			07/18/25 17:00	
			TSS	SM2540 D			07/23/25 09:30	
Q2644-02	RW8-SP303-2025071 7	WATER			07/17/25 13:13			07/18/25
			TDS	SM2540 C			07/18/25 17:00	
			TSS	SM2540 D			07/23/25 09:30	



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/17/25 13:05
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/18/25
Client Sample ID:	RW8-SP100-20250717	SDG No.:	Q2644
Lab Sample ID:	Q2644-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TDS	43.0		1	1.00	10.0	10.0	mg/L		07/18/25 17:00	SM 2540 C-20
TSS	4.00	U	1	1.00	4.00	4.00	mg/L		07/23/25 09:30	SM 2540 D-20

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	07/17/25 13:13
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	07/18/25
Client Sample ID:	RW8-SP303-20250717	SDG No.:	Q2644
Lab Sample ID:	Q2644-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TDS	42.0		1	1.00	10.0	10.0	mg/L		07/18/25 17:00	SM 2540 C-20
TSS	4.00	U	1	1.00	4.00	4.00	mg/L		07/23/25 09:30	SM 2540 D-20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

QC RESULT SUMMARY

Preparation Blank Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2644

Project: NWIRP Bethpage 112G08005-WE13

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136545BL							
TDS	mg/L	< 5.0000	5.0000	U	1.0	10	07/18/2025
Sample ID: LB136573BL							
TSS	mg/L	1	2.0000	J	1	4	07/23/2025

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2644
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q2644-02
Client ID:	RW8-SP303-20250717DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TDS	mg/L	+/-5	42.0		40.0		1	4.88		07/18/2025

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2644
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q2646-01
Client ID:	FRAC TANKDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TSS	mg/L	+/-5	562		559		1	0.54		07/23/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2644

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB136545

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136545BS							
TDS	mg/L	100	95.0		95	1	90-110	07/18/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2644

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB136573

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136573BS							
TSS	mg/L	550	533		97	1	90-110	07/23/2025