

## Duplicate Sample Summary

|                                               |                                           |
|-----------------------------------------------|-------------------------------------------|
| <b>Client:</b> Tetra Tech NUS, Inc.           | <b>SDG No.:</b> Q2644                     |
| <b>Project:</b> NWIRP Bethpage 112G08005-WE13 | <b>Sample ID:</b> Q2644-02                |
| <b>Client ID:</b> RW8-SP303-20250717DUP       | <b>Percent Solids for Spike Sample:</b> 0 |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| TDS     | mg/L  | +/-5                | 42.0             |                    | 40.0                |                    | 1                  | 4.88       |      | 07/18/2025       |

## Duplicate Sample Summary

|                                               |                                           |
|-----------------------------------------------|-------------------------------------------|
| <b>Client:</b> Tetra Tech NUS, Inc.           | <b>SDG No.:</b> Q2644                     |
| <b>Project:</b> NWIRP Bethpage 112G08005-WE13 | <b>Sample ID:</b> Q2646-01                |
| <b>Client ID:</b> FRAC TANKDUP                | <b>Percent Solids for Spike Sample:</b> 0 |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| TSS     | mg/L  | +/-5                | 562              |                    | 559                 |                    | 1                  | 0.54       |      | 07/23/2025       |