

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2644

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4060	4000	102	90 - 110	P	07/22/2025	12:03	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2644

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	107	100	107	80 - 120	P	07/22/2025	12:10	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2644

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5030	5000	101	90 - 110	P	07/22/2025	13:00	LB136571
CCV02	Iron	5020	5000	100	90 - 110	P	07/22/2025	14:05	LB136571
CCV03	Iron	5150	5000	103	90 - 110	P	07/22/2025	15:20	LB136571
CCV04	Iron	5270	5000	105	90 - 110	P	07/22/2025	16:13	LB136571
CCV05	Iron	5300	5000	106	90 - 110	P	07/22/2025	17:14	LB136571
CCV06	Iron	5390	5000	108	90 - 110	P	07/22/2025	18:05	LB136571
CCV07	Iron	5140	5000	103	90 - 110	P	07/22/2025	18:54	LB136571
CCV08	Iron	5340	5000	107	90 - 110	P	07/22/2025	19:19	LB136571