

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tetra Tech NUS, Inc. level: low sdg no.: Q2644
contract: TETR06 lab code: ACE
matrix: Water sample id: Q2644-02 client id: RW8-SP303-20250717MS
Percent Solids for Sample: NA Spiked ID: Q2644-02MS Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1530		54.6		1500	98		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tetra Tech NUS, Inc.</u>	level: <u>low</u>	sdg no.: <u>Q2644</u>
contract: <u>TETR06</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2644-02</u>	client id: <u>RW8-SP303-20250717MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2644-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	1600		54.6		1500	103		P