

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

<b>Client:</b>	<u>Tetra Tech NUS, Inc.</u>	<b>Level:</b>	<u>LOW</u>	<b>SDG No.:</b>	<u>Q2644</u>
<b>Contract:</b>	<u>TETR06</u>			<b>Lab Code:</b>	<u>ACE</u>
<b>Matrix:</b>	<u>Water</u>	<b>Sample ID:</b>	<u>Q2644-02</u>	<b>Client ID:</b>	<u>RW8-SP303-20250717DUP</u>
<b>Percent Solids for Sample:</b>	NA	<b>Duplicate ID</b>	Q2644-02DUP	<b>Percent Solids for Spike Sample:</b>	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	54.6		58.1		6		P

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

<b>Client:</b>	<u>Tetra Tech NUS, Inc.</u>	<b>Level:</b>	<u>LOW</u>	<b>SDG No.:</b>	<u>Q2644</u>
<b>Contract:</b>	<u>TETR06</u>			<b>Lab Code:</b>	<u>ACE</u>
<b>Matrix:</b>	<u>Water</u>	<b>Sample ID:</b>	<u>Q2644-02MS</u>	<b>Client ID:</b>	<u>RW8-SP303-20250717MSD</u>
<b>Percent Solids for Sample:</b>	NA	<b>Duplicate ID</b>	Q2644-02MSD	<b>Percent Solids for Spike Sample:</b>	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Iron	ug/L	20	1530		1600		4		P