

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2644

Client ID : SSTDCCC0.4

Date Analyzed: 07/22/2025

Lab File ID: BN037532.D

Time Analyzed: 10:51

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2751	7.717	7206	10.50	3790	14.36
UPPER LIMIT	5502	8.217	14412	10.998	7580	14.855
LOWER LIMIT	1375.5	7.217	3603	9.998	1895	13.855
EPA SAMPLE NO.						
01 PB168952BL	3218	7.72	8001	10.51	4007	14.36
02 PB168952BS	2046	7.72	5016	10.50	2405	14.36
03 PB168952BSD	1815	7.72	4389	10.50	2110	14.36
04 RW8-SP100-20250717	1735	7.72	4230	10.51	2216	14.36
05 RW8-SP303-20250717	1486	7.72	3623	10.51	1895	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.