

Initial and Continuing Calibration Verification

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Project: NWIRP Bethpage 112G08005-WE13

RunNo.: LB136550

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	07/21/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	07/21/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	07/21/2025