

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV36	Mercury	3.82	4.0	96	90 - 110	CV	07/22/2025	15:13	LB136570

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV34	Mercury	4.50	5.0	90	90 - 110	CV	07/22/2025	15:17	LB136570
CCV35	Mercury	4.91	5.0	98	90 - 110	CV	07/22/2025	15:54	LB136570
CCV36	Mercury	5.27	5.0	105	90 - 110	CV	07/22/2025	16:28	LB136570
CCV37	Mercury	5.14	5.0	103	90 - 110	CV	07/22/2025	16:58	LB136570
CCV38	Mercury	4.98	5.0	100	90 - 110	CV	07/22/2025	17:31	LB136570
CCV39	Mercury	4.96	5.0	99	90 - 110	CV	07/22/2025	18:15	LB136570
CCV40	Mercury	5.05	5.0	101	90 - 110	CV	07/22/2025	18:56	LB136570

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3880	4000	97	90 - 110	P	07/22/2025	12:03	LB136571
	Barium	8140	8000	102	90 - 110	P	07/22/2025	12:03	LB136571
	Cadmium	1950	2000	98	90 - 110	P	07/22/2025	12:03	LB136571
	Chromium	819	800	102	90 - 110	P	07/22/2025	12:03	LB136571
	Lead	3800	4000	95	90 - 110	P	07/22/2025	12:03	LB136571
	Selenium	3880	4000	97	90 - 110	P	07/22/2025	12:03	LB136571
	Silver	1040	1000	104	90 - 110	P	07/22/2025	12:03	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	21.8	20.0	109	80 - 120	P	07/22/2025	12:10	LB136571
	Barium	101	100	102	80 - 120	P	07/22/2025	12:10	LB136571
	Cadmium	5.75	6.0	96	80 - 120	P	07/22/2025	12:10	LB136571
	Chromium	10.8	10.0	108	80 - 120	P	07/22/2025	12:10	LB136571
	Lead	13.2	12.0	110	80 - 120	P	07/22/2025	12:10	LB136571
	Selenium	21.7	20.0	109	80 - 120	P	07/22/2025	12:10	LB136571
	Silver	10.4	10.0	104	80 - 120	P	07/22/2025	12:10	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5010	5000	100	90 - 110	P	07/22/2025	13:00	LB136571
	Barium	10100	10000	101	90 - 110	P	07/22/2025	13:00	LB136571
	Cadmium	2470	2500	99	90 - 110	P	07/22/2025	13:00	LB136571
	Chromium	984	1000	98	90 - 110	P	07/22/2025	13:00	LB136571
	Lead	4920	5000	98	90 - 110	P	07/22/2025	13:00	LB136571
	Selenium	5030	5000	101	90 - 110	P	07/22/2025	13:00	LB136571
	Silver	1260	1250	100	90 - 110	P	07/22/2025	13:00	LB136571
CCV02	Arsenic	5090	5000	102	90 - 110	P	07/22/2025	14:05	LB136571
	Barium	10500	10000	105	90 - 110	P	07/22/2025	14:05	LB136571
	Cadmium	2490	2500	100	90 - 110	P	07/22/2025	14:05	LB136571
	Chromium	976	1000	98	90 - 110	P	07/22/2025	14:05	LB136571
	Lead	4970	5000	99	90 - 110	P	07/22/2025	14:05	LB136571
	Selenium	5100	5000	102	90 - 110	P	07/22/2025	14:05	LB136571
	Silver	1280	1250	102	90 - 110	P	07/22/2025	14:05	LB136571
CCV03	Arsenic	5210	5000	104	90 - 110	P	07/22/2025	15:20	LB136571
	Barium	10600	10000	106	90 - 110	P	07/22/2025	15:20	LB136571
	Cadmium	2560	2500	102	90 - 110	P	07/22/2025	15:20	LB136571
	Chromium	1010	1000	101	90 - 110	P	07/22/2025	15:20	LB136571
	Lead	5110	5000	102	90 - 110	P	07/22/2025	15:20	LB136571
	Selenium	5230	5000	104	90 - 110	P	07/22/2025	15:20	LB136571
	Silver	1310	1250	105	90 - 110	P	07/22/2025	15:20	LB136571
CCV04	Arsenic	5220	5000	104	90 - 110	P	07/22/2025	16:13	LB136571
	Barium	10800	10000	108	90 - 110	P	07/22/2025	16:13	LB136571
	Cadmium	2560	2500	102	90 - 110	P	07/22/2025	16:13	LB136571
	Chromium	1010	1000	101	90 - 110	P	07/22/2025	16:13	LB136571
	Lead	5100	5000	102	90 - 110	P	07/22/2025	16:13	LB136571
	Selenium	5240	5000	105	90 - 110	P	07/22/2025	16:13	LB136571
	Silver	1320	1250	105	90 - 110	P	07/22/2025	16:13	LB136571
CCV05	Arsenic	5170	5000	104	90 - 110	P	07/22/2025	17:14	LB136571
	Barium	10700	10000	107	90 - 110	P	07/22/2025	17:14	LB136571
	Cadmium	2560	2500	102	90 - 110	P	07/22/2025	17:14	LB136571
	Chromium	1020	1000	102	90 - 110	P	07/22/2025	17:14	LB136571
	Lead	5120	5000	102	90 - 110	P	07/22/2025	17:14	LB136571
	Selenium	5180	5000	104	90 - 110	P	07/22/2025	17:14	LB136571

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1320	1250	106	90 - 110	P	07/22/2025	17:14	LB136571
CCV06	Arsenic	5170	5000	103	90 - 110	P	07/22/2025	18:05	LB136571
	Barium	10700	10000	107	90 - 110	P	07/22/2025	18:05	LB136571
	Cadmium	2560	2500	102	90 - 110	P	07/22/2025	18:05	LB136571
	Chromium	1040	1000	104	90 - 110	P	07/22/2025	18:05	LB136571
	Lead	5120	5000	102	90 - 110	P	07/22/2025	18:05	LB136571
	Selenium	5170	5000	103	90 - 110	P	07/22/2025	18:05	LB136571
	Silver	1330	1250	107	90 - 110	P	07/22/2025	18:05	LB136571
CCV07	Arsenic	5190	5000	104	90 - 110	P	07/22/2025	18:54	LB136571
	Barium	10600	10000	106	90 - 110	P	07/22/2025	18:54	LB136571
	Cadmium	2570	2500	103	90 - 110	P	07/22/2025	18:54	LB136571
	Chromium	1010	1000	101	90 - 110	P	07/22/2025	18:54	LB136571
	Lead	5130	5000	103	90 - 110	P	07/22/2025	18:54	LB136571
	Selenium	5190	5000	104	90 - 110	P	07/22/2025	18:54	LB136571
	Silver	1320	1250	105	90 - 110	P	07/22/2025	18:54	LB136571
CCV08	Arsenic	5260	5000	105	90 - 110	P	07/22/2025	19:19	LB136571
	Barium	10800	10000	108	90 - 110	P	07/22/2025	19:19	LB136571
	Cadmium	2580	2500	103	90 - 110	P	07/22/2025	19:19	LB136571
	Chromium	1050	1000	105	90 - 110	P	07/22/2025	19:19	LB136571
	Lead	5150	5000	103	90 - 110	P	07/22/2025	19:19	LB136571
	Selenium	5290	5000	106	90 - 110	P	07/22/2025	19:19	LB136571
	Silver	1330	1250	107	90 - 110	P	07/22/2025	19:19	LB136571