



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB36	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	15:15	LB136570

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB34	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	15:20	LB136570
CCB35	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	15:59	LB136570
CCB36	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	16:30	LB136570
CCB37	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	17:00	LB136570
CCB38	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	17:33	LB136570
CCB39	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	18:20	LB136570
CCB40	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/22/2025	18:58	LB136570

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	15.0	20.0	P	07/22/2025	12:27	LB136571
	Barium	14.6	+/-50	U	25.0	100	P	07/22/2025	12:27	LB136571
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/22/2025	12:27	LB136571
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/22/2025	12:27	LB136571
	Lead	2.30	+/-6	U	9.60	12.0	P	07/22/2025	12:27	LB136571
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/22/2025	12:27	LB136571
	Silver	1.62	+/-5	U	5.00	10.0	P	07/22/2025	12:27	LB136571

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	15.0	20.0	P	07/22/2025	13:05	LB136571
	Barium	14.6	+/-50	U	25.0	100	P	07/22/2025	13:05	LB136571
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/22/2025	13:05	LB136571
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/22/2025	13:05	LB136571
	Lead	2.30	+/-6	U	9.60	12.0	P	07/22/2025	13:05	LB136571
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/22/2025	13:05	LB136571
	Silver	1.62	+/-5	U	5.00	10.0	P	07/22/2025	13:05	LB136571
CCB02	Arsenic	5.12	+/-10	U	15.0	20.0	P	07/22/2025	14:09	LB136571
	Barium	14.6	+/-50	U	25.0	100	P	07/22/2025	14:09	LB136571
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/22/2025	14:09	LB136571
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/22/2025	14:09	LB136571
	Lead	2.30	+/-6	U	9.60	12.0	P	07/22/2025	14:09	LB136571
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/22/2025	14:09	LB136571
	Silver	1.62	+/-5	U	5.00	10.0	P	07/22/2025	14:09	LB136571
CCB03	Arsenic	5.12	+/-10	U	15.0	20.0	P	07/22/2025	15:26	LB136571
	Barium	14.6	+/-50	U	25.0	100	P	07/22/2025	15:26	LB136571
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/22/2025	15:26	LB136571
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/22/2025	15:26	LB136571
	Lead	2.30	+/-6	U	9.60	12.0	P	07/22/2025	15:26	LB136571
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/22/2025	15:26	LB136571
	Silver	1.62	+/-5	U	5.00	10.0	P	07/22/2025	15:26	LB136571
CCB04	Arsenic	5.12	+/-10	U	15.0	20.0	P	07/22/2025	16:17	LB136571
	Barium	14.6	+/-50	U	25.0	100	P	07/22/2025	16:17	LB136571
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/22/2025	16:17	LB136571
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/22/2025	16:17	LB136571
	Lead	2.30	+/-6	U	9.60	12.0	P	07/22/2025	16:17	LB136571
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/22/2025	16:17	LB136571
	Silver	1.62	+/-5	U	5.00	10.0	P	07/22/2025	16:17	LB136571
CCB05	Arsenic	5.12	+/-10	U	15.0	20.0	P	07/22/2025	17:18	LB136571
	Barium	14.6	+/-50	U	25.0	100	P	07/22/2025	17:18	LB136571
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/22/2025	17:18	LB136571
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/22/2025	17:18	LB136571
	Lead	2.30	+/-6	U	9.60	12.0	P	07/22/2025	17:18	LB136571
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/22/2025	17:18	LB136571
	Silver	1.62	+/-5	U	5.00	10.0	P	07/22/2025	17:18	LB136571
CCB06	Arsenic	5.12	+/-10	U	15.0	20.0	P	07/22/2025	18:09	LB136571
	Barium	14.6	+/-50	U	25.0	100	P	07/22/2025	18:09	LB136571
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/22/2025	18:09	LB136571
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/22/2025	18:09	LB136571

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q2645

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	9.60	12.0	P	07/22/2025	18:09	LB136571
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/22/2025	18:09	LB136571
	Silver	1.62	+/-5	U	5.00	10.0	P	07/22/2025	18:09	LB136571
CCB07	Arsenic	5.12	+/-10	U	15.0	20.0	P	07/22/2025	18:58	LB136571
	Barium	14.6	+/-50	U	25.0	100	P	07/22/2025	18:58	LB136571
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/22/2025	18:58	LB136571
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/22/2025	18:58	LB136571
	Lead	2.30	+/-6	U	9.60	12.0	P	07/22/2025	18:58	LB136571
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/22/2025	18:58	LB136571
	Silver	1.62	+/-5	U	5.00	10.0	P	07/22/2025	18:58	LB136571
CCB08	Arsenic	5.12	+/-10	U	15.0	20.0	P	07/22/2025	19:23	LB136571
	Barium	14.6	+/-50	U	25.0	100	P	07/22/2025	19:23	LB136571
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/22/2025	19:23	LB136571
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/22/2025	19:23	LB136571
	Lead	2.30	+/-6	U	9.60	12.0	P	07/22/2025	19:23	LB136571
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/22/2025	19:23	LB136571
	Silver	1.62	+/-5	U	5.00	10.0	P	07/22/2025	19:23	LB136571