

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Tetra Tech NUS, Inc.</u>	SDG No.: <u>Q2645</u>
Contract: <u>TETR06</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	10.7			-20	20	07/22/2025	12:37	LB136571
	Barium	2.26	6.0	38	-94	106	07/22/2025	12:37	LB136571
	Cadmium	-1.25	1.0	125	-5	7	07/22/2025	12:37	LB136571
	Chromium	55.9	52.0	108	42	62	07/22/2025	12:37	LB136571
	Lead	-3.81			-12	12	07/22/2025	12:37	LB136571
	Selenium	-9.01			-20	20	07/22/2025	12:37	LB136571
	Silver	3.81			-10	10	07/22/2025	12:37	LB136571
ICSAB01	Arsenic	103	104	99	88.4	120	07/22/2025	12:48	LB136571
	Barium	494	537	92	437	637	07/22/2025	12:48	LB136571
	Cadmium	914	972	94	826	1120	07/22/2025	12:48	LB136571
	Chromium	536	542	99	460	624	07/22/2025	12:48	LB136571
	Lead	42.1	49.0	86	37	61	07/22/2025	12:48	LB136571
	Selenium	40.7	46.0	88	26	66	07/22/2025	12:48	LB136571
	Silver	210	201	104	170	232	07/22/2025	12:48	LB136571