

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Continuing Calib Date: 07/22/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 00:34

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.76	4.76	4.66	4.86	0.00
Aroclor-1248-2	(2)	5.00	5.00	4.90	5.10	0.01
Aroclor-1248-3	(3)	5.21	5.21	5.11	5.31	0.00
Aroclor-1248-4	(4)	5.56	5.57	5.47	5.67	0.01
Aroclor-1248-5	(5)	5.60	5.61	5.51	5.71	0.01
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.70	8.70	8.60	8.80	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Continuing Calib Date: 07/22/2025

Initial Calibration Date(s): 07/08/2025

07/08/2025

Continuing Calib Time: 00:34

Initial Calibration Time(s): 14:06

22:16

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.74	4.74	4.64	4.84	0.00
Aroclor-1248-2 (2)	4.98	4.98	4.88	5.08	0.01
Aroclor-1248-3 (3)	5.02	5.02	4.92	5.12	0.00
Aroclor-1248-4 (4)	5.19	5.19	5.09	5.29	0.00
Aroclor-1248-5 (5)	5.58	5.58	5.48	5.68	0.00
Tetrachloro-m-xylene	3.66	3.67	3.57	3.77	0.01
Decachlorobiphenyl	8.64	8.65	8.55	8.75	0.01