

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Continuing Calib Date: 07/21/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 15:18

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.63	5.64	5.54	5.74	0.01
Aroclor-1016-2	(2)	5.66	5.66	5.56	5.76	0.00
Aroclor-1016-3	(3)	5.72	5.72	5.62	5.82	0.00
Aroclor-1016-4	(4)	5.82	5.82	5.72	5.92	0.00
Aroclor-1016-5	(5)	6.11	6.11	6.01	6.21	0.00
Aroclor-1260-1	(1)	7.22	7.23	7.13	7.33	0.01
Aroclor-1260-2	(2)	7.48	7.48	7.38	7.58	0.00
Aroclor-1260-3	(3)	7.84	7.84	7.74	7.94	0.00
Aroclor-1260-4	(4)	8.06	8.06	7.96	8.16	0.00
Aroclor-1260-5	(5)	8.38	8.38	8.28	8.48	0.00
Tetrachloro-m-xylene		4.48	4.49	4.39	4.59	0.01
Decachlorobiphenyl		10.17	10.18	10.08	10.28	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Continuing Calib Date: 07/21/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 15:18

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.85	4.86	4.76	4.96	0.01
Aroclor-1016-2	(2)	4.87	4.88	4.78	4.98	0.01
Aroclor-1016-3	(3)	5.05	5.05	4.95	5.15	0.00
Aroclor-1016-4	(4)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-5	(5)	5.30	5.31	5.21	5.41	0.01
Aroclor-1260-1	(1)	6.33	6.34	6.24	6.44	0.01
Aroclor-1260-2	(2)	6.52	6.53	6.43	6.63	0.01
Aroclor-1260-3	(3)	6.67	6.68	6.58	6.78	0.01
Aroclor-1260-4	(4)	7.14	7.15	7.05	7.25	0.01
Aroclor-1260-5	(5)	7.38	7.39	7.29	7.49	0.01
Tetrachloro-m-xylene		3.77	3.78	3.68	3.88	0.01
Decachlorobiphenyl		8.77	8.78	8.68	8.88	0.01