

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Continuing Calib Date: 07/21/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 20:45

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.63	5.64	5.54	5.74	0.01
Aroclor-1242-2 (2)	5.65	5.66	5.56	5.76	0.01
Aroclor-1242-3 (3)	5.71	5.72	5.62	5.82	0.01
Aroclor-1242-4 (4)	5.81	5.82	5.72	5.92	0.01
Aroclor-1242-5 (5)	6.54	6.55	6.45	6.65	0.01
Tetrachloro-m-xylene	4.48	4.49	4.39	4.59	0.01
Decachlorobiphenyl	10.16	10.17	10.07	10.27	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Continuing Calib Date: 07/21/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 20:45

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.85	4.86	4.76	4.96	0.01
Aroclor-1242-2 (2)	4.87	4.88	4.78	4.98	0.01
Aroclor-1242-3 (3)	5.04	5.05	4.95	5.15	0.01
Aroclor-1242-4 (4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1242-5 (5)	5.65	5.66	5.56	5.76	0.01
Tetrachloro-m-xylene	3.77	3.78	3.68	3.88	0.01
Decachlorobiphenyl	8.77	8.78	8.68	8.88	0.01