

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Continuing Calib Date: 07/22/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 02:11

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.64       | 5.64      | 5.54      | 5.74  | 0.00       |
| Aroclor-1248-2 (2)   | 5.91       | 5.91      | 5.81      | 6.01  | 0.00       |
| Aroclor-1248-3 (3)   | 6.11       | 6.11      | 6.01      | 6.21  | 0.00       |
| Aroclor-1248-4 (4)   | 6.51       | 6.51      | 6.41      | 6.61  | 0.00       |
| Aroclor-1248-5 (5)   | 6.55       | 6.55      | 6.45      | 6.65  | 0.00       |
| Tetrachloro-m-xylene | 4.48       | 4.49      | 4.39      | 4.59  | 0.01       |
| Decachlorobiphenyl   | 10.17      | 10.17     | 10.07     | 10.27 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Continuing Calib Date: 07/22/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 02:11

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.85       | 4.86      | 4.76      | 4.96 | 0.01       |
| Aroclor-1248-2 (2)   | 5.09       | 5.09      | 4.99      | 5.19 | 0.00       |
| Aroclor-1248-3 (3)   | 5.13       | 5.14      | 5.04      | 5.24 | 0.01       |
| Aroclor-1248-4 (4)   | 5.30       | 5.31      | 5.21      | 5.41 | 0.01       |
| Aroclor-1248-5 (5)   | 5.69       | 5.70      | 5.60      | 5.80 | 0.01       |
| Tetrachloro-m-xylene | 3.77       | 3.78      | 3.68      | 3.88 | 0.01       |
| Decachlorobiphenyl   | 8.77       | 8.78      | 8.68      | 8.88 | 0.01       |