

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** TETR06  
**Lab Code:** ACE **SDG NO.:** Q2645  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 07/22/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO112366.D **Time Analyzed:** 08:57

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.759 | 4.661     | 4.861 | 518.970            | 500.000           | 3.8  |
| Aroclor-1016-2       | 4.778 | 4.680     | 4.880 | 532.520            | 500.000           | 6.5  |
| Aroclor-1016-3       | 4.835 | 4.736     | 4.936 | 514.470            | 500.000           | 2.9  |
| Aroclor-1016-4       | 4.954 | 4.856     | 5.056 | 522.410            | 500.000           | 4.5  |
| Aroclor-1016-5       | 5.211 | 5.113     | 5.313 | 511.450            | 500.000           | 2.3  |
| Aroclor-1260-1       | 6.249 | 6.151     | 6.351 | 493.780            | 500.000           | -1.2 |
| Aroclor-1260-2       | 6.439 | 6.341     | 6.541 | 495.000            | 500.000           | -1.0 |
| Aroclor-1260-3       | 6.805 | 6.707     | 6.907 | 500.210            | 500.000           | 0.0  |
| Aroclor-1260-4       | 7.064 | 6.967     | 7.167 | 503.510            | 500.000           | 0.7  |
| Aroclor-1260-5       | 7.307 | 7.209     | 7.409 | 486.250            | 500.000           | -2.8 |
| Decachlorobiphenyl   | 8.700 | 8.603     | 8.803 | 49.980             | 50.000            | 0.0  |
| Tetrachloro-m-xylene | 3.672 | 3.573     | 3.773 | 57.200             | 50.000            | 14.4 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** TETR06  
**Lab Code:** ACE **SDG NO.:** Q2645  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 07/22/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO112366.D **Time Analyzed:** 08:57

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.737 | 4.661     | 4.861 | 471.470            | 500.000           | -5.7 |
| Aroclor-1016-2       | 4.756 | 4.679     | 4.879 | 469.570            | 500.000           | -6.1 |
| Aroclor-1016-3       | 4.930 | 4.854     | 5.054 | 484.320            | 500.000           | -3.1 |
| Aroclor-1016-4       | 4.973 | 4.897     | 5.097 | 496.800            | 500.000           | -0.6 |
| Aroclor-1016-5       | 5.185 | 5.109     | 5.309 | 472.560            | 500.000           | -5.5 |
| Aroclor-1260-1       | 6.214 | 6.138     | 6.338 | 482.800            | 500.000           | -3.4 |
| Aroclor-1260-2       | 6.403 | 6.326     | 6.526 | 514.630            | 500.000           | 2.9  |
| Aroclor-1260-3       | 6.554 | 6.478     | 6.678 | 517.520            | 500.000           | 3.5  |
| Aroclor-1260-4       | 7.024 | 6.948     | 7.148 | 501.240            | 500.000           | 0.2  |
| Aroclor-1260-5       | 7.267 | 7.190     | 7.390 | 519.750            | 500.000           | 4.0  |
| Decachlorobiphenyl   | 8.643 | 8.568     | 8.768 | 47.160             | 50.000            | -5.7 |
| Tetrachloro-m-xylene | 3.662 | 3.584     | 3.784 | 51.060             | 50.000            | 2.1  |