

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q2645
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073972.D **Time Analyzed:** 08:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.633	5.539	5.739	451.720	500.000	-9.7
Aroclor-1016-2	5.654	5.560	5.760	488.530	500.000	-2.3
Aroclor-1016-3	5.716	5.622	5.822	479.600	500.000	-4.1
Aroclor-1016-4	5.814	5.720	5.920	483.450	500.000	-3.3
Aroclor-1016-5	6.106	6.012	6.212	488.110	500.000	-2.4
Aroclor-1260-1	7.223	7.129	7.329	470.410	500.000	-5.9
Aroclor-1260-2	7.476	7.383	7.583	429.810	500.000	-14.0
Aroclor-1260-3	7.834	7.740	7.940	494.290	500.000	-1.1
Aroclor-1260-4	8.057	7.964	8.164	508.080	500.000	1.6
Aroclor-1260-5	8.375	8.283	8.483	513.090	500.000	2.6
Decachlorobiphenyl	10.166	10.076	10.276	53.820	50.000	7.6
Tetrachloro-m-xylene	4.482	4.388	4.588	51.200	50.000	2.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q2645
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073972.D **Time Analyzed:** 08:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.849	4.758	4.958	471.860	500.000	-5.6
Aroclor-1016-2	4.867	4.775	4.975	478.870	500.000	-4.2
Aroclor-1016-3	5.043	4.951	5.151	489.440	500.000	-2.1
Aroclor-1016-4	5.085	4.993	5.193	487.850	500.000	-2.4
Aroclor-1016-5	5.298	5.207	5.407	494.440	500.000	-1.1
Aroclor-1260-1	6.328	6.237	6.437	472.030	500.000	-5.6
Aroclor-1260-2	6.516	6.425	6.625	470.640	500.000	-5.9
Aroclor-1260-3	6.668	6.577	6.777	469.520	500.000	-6.1
Aroclor-1260-4	7.137	7.046	7.246	489.920	500.000	-2.0
Aroclor-1260-5	7.379	7.289	7.489	491.630	500.000	-1.7
Decachlorobiphenyl	8.769	8.680	8.880	55.590	50.000	11.2
Tetrachloro-m-xylene	3.772	3.678	3.878	48.690	50.000	-2.6