

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q2645
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073987.D **Time Analyzed:** 15:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.634	5.539	5.739	447.980	500.000	-10.4
Aroclor-1016-2	5.655	5.560	5.760	488.950	500.000	-2.2
Aroclor-1016-3	5.717	5.622	5.822	483.620	500.000	-3.3
Aroclor-1016-4	5.815	5.720	5.920	487.720	500.000	-2.5
Aroclor-1016-5	6.107	6.012	6.212	490.820	500.000	-1.8
Aroclor-1260-1	7.223	7.129	7.329	477.040	500.000	-4.6
Aroclor-1260-2	7.477	7.383	7.583	431.310	500.000	-13.7
Aroclor-1260-3	7.835	7.740	7.940	484.320	500.000	-3.1
Aroclor-1260-4	8.058	7.964	8.164	499.440	500.000	-0.1
Aroclor-1260-5	8.376	8.283	8.483	502.480	500.000	0.5
Decachlorobiphenyl	10.168	10.076	10.276	53.680	50.000	7.4
Tetrachloro-m-xylene	4.482	4.388	4.588	50.840	50.000	1.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q2645
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073987.D **Time Analyzed:** 15:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.851	4.758	4.958	523.290	500.000	4.7
Aroclor-1016-2	4.869	4.775	4.975	525.770	500.000	5.2
Aroclor-1016-3	5.045	4.951	5.151	530.150	500.000	6.0
Aroclor-1016-4	5.086	4.993	5.193	529.020	500.000	5.8
Aroclor-1016-5	5.301	5.207	5.407	537.530	500.000	7.5
Aroclor-1260-1	6.330	6.237	6.437	529.710	500.000	5.9
Aroclor-1260-2	6.519	6.425	6.625	534.970	500.000	7.0
Aroclor-1260-3	6.670	6.577	6.777	517.150	500.000	3.4
Aroclor-1260-4	7.139	7.046	7.246	539.350	500.000	7.9
Aroclor-1260-5	7.382	7.289	7.489	536.370	500.000	7.3
Decachlorobiphenyl	8.772	8.680	8.880	60.720	50.000	21.4
Tetrachloro-m-xylene	3.774	3.678	3.878	53.660	50.000	7.3