

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070825

Review By	yogesh	Review On	7/8/2025 1:10:21 PM
Supervise By	mohammad	Supervise On	7/10/2025 2:03:44 AM
SubDirectory	PO070825	HP Acquire Method	HP Processing Method PO070825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112086.D	08 Jul 2025 13:32		YP/AJ	Ok
2	I.BLK	I.BLK	PO112087.D	08 Jul 2025 13:49		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO112088.D	08 Jul 2025 14:06		YP/AJ	Ok,M
4	AR1660ICC750	AR1660ICC750	PO112089.D	08 Jul 2025 14:25		YP/AJ	Ok,M
5	AR1660ICC500	AR1660ICC500	PO112090.D	08 Jul 2025 14:43		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO112091.D	08 Jul 2025 15:02		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO112092.D	08 Jul 2025 15:21		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO112093.D	08 Jul 2025 15:38		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO112094.D	08 Jul 2025 15:57		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO112095.D	08 Jul 2025 16:15		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO112096.D	08 Jul 2025 16:34		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO112097.D	08 Jul 2025 16:52		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO112098.D	08 Jul 2025 17:11		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO112099.D	08 Jul 2025 17:29		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO112100.D	08 Jul 2025 17:48		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO112101.D	08 Jul 2025 18:05		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO112102.D	08 Jul 2025 18:24		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO112103.D	08 Jul 2025 18:41		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO070825

Review By	yogesh	Review On	7/8/2025 1:10:21 PM
Supervise By	mohammad	Supervise On	7/10/2025 2:03:44 AM
SubDirectory	PO070825	HP Acquire Method	HP Processing Method PO070825
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PO112104.D	08 Jul 2025 18:59		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO112105.D	08 Jul 2025 19:18		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO112106.D	08 Jul 2025 19:35		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO112107.D	08 Jul 2025 19:54		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO112108.D	08 Jul 2025 20:11		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO112109.D	08 Jul 2025 20:28		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO112110.D	08 Jul 2025 20:47		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO112111.D	08 Jul 2025 21:04		YP/AJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PO112112.D	08 Jul 2025 21:22		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO112113.D	08 Jul 2025 21:39		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO112114.D	08 Jul 2025 21:57		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO112115.D	08 Jul 2025 22:16		YP/AJ	Ok,M
31	PO070825ICV500	ICVPO070825	PO112116.D	08 Jul 2025 22:34		YP/AJ	Ok,M
32	AR1242ICV500	ICVPO070825AR1242	PO112117.D	08 Jul 2025 23:11		YP/AJ	Ok
33	AR1248ICV500	ICVPO070825AR1248	PO112118.D	08 Jul 2025 23:47		YP/AJ	Ok
34	AR1254ICV500	ICVPO070825AR1254	PO112119.D	09 Jul 2025 00:24		YP/AJ	Ok
35	AR1268ICV500	ICVPO070825AR1268	PO112120.D	09 Jul 2025 01:01		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072125

Review By	yogesh	Review On	7/21/2025 1:01:47 PM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:05 AM
SubDirectory	PO072125	HP Acquire Method	HP Processing Method PO070825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112331.D	21 Jul 2025 08:47		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112332.D	21 Jul 2025 10:15		YP/AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PO112333.D	21 Jul 2025 10:41	TCMX high in first column	YP/AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PO112334.D	21 Jul 2025 11:01		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO112335.D	21 Jul 2025 11:38		YP/AJ	Ok
6	I.BLK	I.BLK	PO112336.D	21 Jul 2025 12:34		YP/AJ	Ok
7	DDT ANALOG	DDT ANALOG	PO112337.D	21 Jul 2025 12:53		YP/AJ	Ok
8	Q2639-03	OU4-TS-39-071725	PO112338.D	21 Jul 2025 13:11		YP/AJ	Ok
9	Q2639-05	OU4-TS-40-071725	PO112339.D	21 Jul 2025 13:30		YP/AJ	Ok
10	Q2639-07	OU4-TS-41-071725	PO112340.D	21 Jul 2025 13:47		YP/AJ	Ok,M
11	Q2639-09	OU4-TS-42-071725	PO112341.D	21 Jul 2025 14:06	DCB Low in both column	YP/AJ	Not Ok
12	Q2639-11	OU4-TS-43-071725	PO112342.D	21 Jul 2025 14:24	DCB Low in both column	YP/AJ	Not Ok
13	Q2639-13	OU4-TS-44-071725	PO112343.D	21 Jul 2025 14:41	DCB Low in both column	YP/AJ	Not Ok
14	Q2641-01	P001-CONCRETE001-	PO112344.D	21 Jul 2025 14:59	AR1254+1260 Hit	YP/AJ	Ok,M
15	Q2645-02	RW5B-CARBON-20250	PO112345.D	21 Jul 2025 15:16	All Surrogate fail	YP/AJ	ReRun
16	AR1660CCC500	AR1660CCC500	PO112346.D	21 Jul 2025 16:10		YP/AJ	Ok,M
17	AR1242CCC500	AR1242CCC500	PO112347.D	21 Jul 2025 16:45		YP/AJ	Ok,M
18	AR1248CCC500	AR1248CCC500	PO112348.D	21 Jul 2025 17:04		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072125

Review By	yogesh	Review On	7/21/2025 1:01:47 PM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:05 AM
SubDirectory	PO072125	HP Acquire Method	HP Processing Method PO070825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1254CCC500	AR1254CCC500	PO112349.D	21 Jul 2025 17:21		YP/AJ	Ok
20	I.BLK	I.BLK	PO112350.D	21 Jul 2025 17:38		YP/AJ	Ok
21	PB168943BL	PB168943BL	PO112351.D	21 Jul 2025 19:42		YP/AJ	Ok
22	PB168943BS	PB168943BS	PO112352.D	21 Jul 2025 20:01		YP/AJ	Ok,M
23	Q2661-01	245F64-1-1	PO112353.D	21 Jul 2025 20:18		YP/AJ	Ok,M
24	Q2661-02	245F64-1-2	PO112354.D	21 Jul 2025 20:37		YP/AJ	Ok,M
25	Q2662-01	L10-A	PO112355.D	21 Jul 2025 20:55	AR1254 Hit	YP/AJ	Ok,M
26	Q2663-01	BC289086-1-1	PO112356.D	21 Jul 2025 21:13		YP/AJ	Ok,M
27	Q2663-02	BC289086-1-2	PO112357.D	21 Jul 2025 21:31		YP/AJ	Ok,M
28	Q2663-03	BC289086-2-1	PO112358.D	21 Jul 2025 21:49	AR1260 Hit	YP/AJ	Ok,M
29	Q2663-04	BC289086-2-2	PO112359.D	21 Jul 2025 22:07	AR1260 Hit	YP/AJ	Ok,M
30	AR1660CCC500	AR1660CCC500	PO112360.D	21 Jul 2025 23:21	DCB low in first column	YP/AJ	Ok,M
31	AR1242CCC500	AR1242CCC500	PO112361.D	22 Jul 2025 00:16	CCC fail on higher side	YP/AJ	Not Ok
32	AR1248CCC500	AR1248CCC500	PO112362.D	22 Jul 2025 00:34		YP/AJ	Ok
33	AR1254CCC500	AR1254CCC500	PO112363.D	22 Jul 2025 00:52	TCMX high in first column	YP/AJ	Ok
34	I.BLK	I.BLK	PO112364.D	22 Jul 2025 01:10		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072225

Review By	yogesh	Review On	7/22/2025 12:28:05 PM
Supervise By	mohammad	Supervise On	7/24/2025 1:25:37 AM
SubDirectory	PO072225	HP Acquire Method	HP Processing Method PO070825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO112365.D	22 Jul 2025 08:18		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO112366.D	22 Jul 2025 08:57		YP/AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PO112367.D	22 Jul 2025 09:15		YP/AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PO112368.D	22 Jul 2025 09:32		YP/AJ	Ok,M
5	AR1254CCC500	AR1254CCC500	PO112369.D	22 Jul 2025 09:49		YP/AJ	Ok,M
6	I.BLK	I.BLK	PO112370.D	22 Jul 2025 10:08		YP/AJ	Ok
7	DDT ANALOG	DDT ANALOG	PO112371.D	22 Jul 2025 10:26		YP/AJ	Ok
8	Q2639-09	OU4-TS-42-071725	PO112372.D	22 Jul 2025 11:08		YP/AJ	Ok
9	Q2639-11	OU4-TS-43-071725	PO112373.D	22 Jul 2025 11:26		YP/AJ	Ok,M
10	Q2639-13	OU4-TS-44-071725	PO112374.D	22 Jul 2025 11:44		YP/AJ	Ok
11	PB168946BL	PB168946BL	PO112375.D	22 Jul 2025 13:35		YP/AJ	Ok
12	PB168946BS	PB168946BS	PO112376.D	22 Jul 2025 13:53		YP/AJ	Ok,M
13	Q2645-02RX	RW5B-CARBON-20250	PO112377.D	22 Jul 2025 14:12	All surrogate fail	YP/AJ	Confirms
14	Q2649-01	WC-1	PO112378.D	22 Jul 2025 14:30		YP/AJ	Ok
15	Q2649-01MS	WC-1MS	PO112379.D	22 Jul 2025 14:48		YP/AJ	Ok,M
16	Q2649-01MSD	WC-1MSD	PO112380.D	22 Jul 2025 15:07	RPD Fail	YP/AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PO112381.D	22 Jul 2025 16:21		YP/AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PO112382.D	22 Jul 2025 16:39		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072225

Review By	yogesh	Review On	7/22/2025 12:28:05 PM
Supervise By	mohammad	Supervise On	7/24/2025 1:25:37 AM
SubDirectory	PO072225	HP Acquire Method	HP Processing Method PO070825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO112383.D	22 Jul 2025 16:56		YP/AJ	Ok,M
20	AR1254CCC500	AR1254CCC500	PO112384.D	22 Jul 2025 17:14		YP/AJ	Ok,M
21	I.BLK	I.BLK	PO112385.D	22 Jul 2025 17:32		YP/AJ	Ok
22	Q2670-01	RBR251375-1	PO112386.D	22 Jul 2025 17:49	AR1248 Hit	YP/AJ	Ok,M
23	Q2670-02	RBR251375-2	PO112387.D	22 Jul 2025 18:08	AR1248 Hit	YP/AJ	Ok,M
24	Q2670-03	72-11995-1	PO112388.D	22 Jul 2025 18:25	AR1248 Hit	YP/AJ	Ok,M
25	Q2670-04	72-11995-2	PO112389.D	22 Jul 2025 18:44	AR1248 Hit	YP/AJ	Ok,M
26	Q2670-05	72-12000-1	PO112390.D	22 Jul 2025 19:02	AR1248 Hit	YP/AJ	Ok,M
27	Q2670-06	72-12000-2	PO112391.D	22 Jul 2025 19:20	AR1248 Hit	YP/AJ	Ok,M
28	Q2670-07	72-11991-1	PO112392.D	22 Jul 2025 19:39	AR1248 Hit	YP/AJ	Ok,M
29	Q2670-08	72-11991-2	PO112393.D	22 Jul 2025 19:57	AR1248 Hit	YP/AJ	Ok,M
30	PB168958BL	PB168958BL	PO112394.D	22 Jul 2025 20:34		YP/AJ	Ok,M
31	PB168958BS	PB168958BS	PO112395.D	22 Jul 2025 20:52		YP/AJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PO112396.D	22 Jul 2025 22:05		YP/AJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PO112397.D	22 Jul 2025 23:00		YP/AJ	Ok,M
34	AR1248CCC500	AR1248CCC500	PO112398.D	22 Jul 2025 23:18		YP/AJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PO112399.D	22 Jul 2025 23:37		YP/AJ	Ok,M
36	I.BLK	I.BLK	PO112400.D	22 Jul 2025 23:55		YP/AJ	Ok,M
37	PB168957BL	PB168957BL	PO112401.D	23 Jul 2025 00:14		YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072225

Review By	yogesh	Review On	7/22/2025 12:28:05 PM
Supervise By	mohammad	Supervise On	7/24/2025 1:25:37 AM
SubDirectory	PO072225	HP Acquire Method	HP Processing Method PO070825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB168957BS	PB168957BS	PO112402.D	23 Jul 2025 00:32		YP/AJ	Ok,M
39	PB168957BSD	PB168957BSD	PO112403.D	23 Jul 2025 00:50		YP/AJ	Ok,M
40	Q2640-01	ELMORA-WATER	PO112404.D	23 Jul 2025 01:09		YP/AJ	Ok,M
41	Q2646-01	FRAC TANK	PO112405.D	23 Jul 2025 01:27		YP/AJ	Ok,M
42	Q2655-03	71125	PO112406.D	23 Jul 2025 01:46		YP/AJ	Ok,M
43	Q2655-04	71425	PO112407.D	23 Jul 2025 02:04		YP/AJ	Ok
44	Q2657-01	12825-A	PO112408.D	23 Jul 2025 02:22		YP/AJ	Ok,M
45	Q2660-05	MOO-25-0220	PO112409.D	23 Jul 2025 02:41		YP/AJ	Ok,M
46	AR1660CCC500	AR1660CCC500	PO112410.D	23 Jul 2025 04:12	DCB low in both column	YP/AJ	Ok,M
47	I.BLK	I.BLK	PO112411.D	23 Jul 2025 05:43		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070825

Review By	yogesh	Review On	7/7/2025 4:17:13 PM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:14 AM
SubDirectory	PP070825	HP Acquire Method	HP Processing Method PP070825

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073552.D	07 Jul 2025 20:13		YP\AJ	Ok
2	I.BLK	I.BLK	PP073553.D	07 Jul 2025 20:30		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP073554.D	07 Jul 2025 21:03		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP073555.D	07 Jul 2025 21:19		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP073556.D	07 Jul 2025 21:35		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP073557.D	07 Jul 2025 21:52		YP\AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP073558.D	07 Jul 2025 22:08		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP073559.D	07 Jul 2025 22:24		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP073560.D	07 Jul 2025 22:41		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP073561.D	07 Jul 2025 22:57		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP073562.D	07 Jul 2025 23:14		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PP073563.D	07 Jul 2025 23:30		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP073564.D	07 Jul 2025 23:46		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PP073565.D	08 Jul 2025 00:03		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PP073566.D	08 Jul 2025 00:19		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP073567.D	08 Jul 2025 00:35		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP073568.D	08 Jul 2025 00:52		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP073569.D	08 Jul 2025 01:08		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070825

Review By	yogesh	Review On	7/7/2025 4:17:13 PM
Supervise By	mohammad	Supervise On	7/9/2025 1:51:14 AM
SubDirectory	PP070825	HP Acquire Method	HP Processing Method PP070825
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PP073570.D	08 Jul 2025 01:25		YPIAJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PP073571.D	08 Jul 2025 01:41		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP073572.D	08 Jul 2025 01:57		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PP073573.D	08 Jul 2025 02:14		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP073574.D	08 Jul 2025 02:30		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP073575.D	08 Jul 2025 02:46		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP073576.D	08 Jul 2025 03:03		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP073577.D	08 Jul 2025 03:19		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP073578.D	08 Jul 2025 03:35		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP073579.D	08 Jul 2025 03:52		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP073580.D	08 Jul 2025 04:08		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP073581.D	08 Jul 2025 04:24		YPIAJ	Ok,M
31	PP070825ICV500	ICVPP070825	PP073582.D	08 Jul 2025 04:41		YPIAJ	Ok
32	AR1242ICV500	ICVPP070825AR1242	PP073583.D	08 Jul 2025 05:30		YPIAJ	Ok
33	AR1248ICV500	ICVPP070825AR1248	PP073584.D	08 Jul 2025 05:46		YPIAJ	Ok
34	AR1254ICV500	ICVPP070825AR1254	PP073585.D	08 Jul 2025 06:19		YPIAJ	Ok
35	AR1268ICV500	ICVPP070825AR1268	PP073586.D	08 Jul 2025 06:52		YPIAJ	Ok
36	DDT ANALOGUE	DDT ANALOGUE	PP073587.D	08 Jul 2025 07:24		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP072125

Review By	yogesh	Review On	7/21/2025 11:07:59 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:32:53 AM
SubDirectory	PP072125	HP Acquire Method	HP Processing Method PP070825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP073971.D	21 Jul 2025 08:35		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP073972.D	21 Jul 2025 08:51		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP073973.D	21 Jul 2025 09:08		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP073974.D	21 Jul 2025 09:24		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP073975.D	21 Jul 2025 09:40		YPIAJ	Ok
6	I.BLK	I.BLK	PP073976.D	21 Jul 2025 09:56		YPIAJ	Ok
7	DDT ANALOG	DDT ANALOG	PP073977.D	21 Jul 2025 10:13		YPIAJ	Ok
8	PB168911BS	PB168911BS	PP073978.D	21 Jul 2025 10:46		YPIAJ	Ok
9	PB168911BSD	PB168911BSD	PP073979.D	21 Jul 2025 11:05		YPIAJ	Ok
10	Q2610-01DL	2010DL	PP073980.D	21 Jul 2025 11:21	AR1242 Hit	YPIAJ	Ok,M
11	Q2610-03DL	2012DL	PP073981.D	21 Jul 2025 11:37	AR1248 Hit	YPIAJ	Ok,M
12	Q2610-05DL	2014DL	PP073982.D	21 Jul 2025 13:07	AR1248 Hit	YPIAJ	Ok,M
13	Q2635-01	RT2286	PP073983.D	21 Jul 2025 13:24	AR1260 Hit	YPIAJ	Ok,M
14	Q2635-01MS	RT2286MS	PP073984.D	21 Jul 2025 13:40		YPIAJ	Ok,M
15	Q2635-01MSD	RT2286MSD	PP073985.D	21 Jul 2025 13:56		YPIAJ	Ok,M
16	Q2634-01	J2315	PP073986.D	21 Jul 2025 14:13		YPIAJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP073987.D	21 Jul 2025 15:18	DCB HIGH in second column	YPIAJ	Ok
18	AR1242CCC500	AR1242CCC500	PP073988.D	21 Jul 2025 15:34		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP072125

Review By	yogesh	Review On	7/21/2025 11:07:59 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:32:53 AM
SubDirectory	PP072125	HP Acquire Method	HP Processing Method PP070825
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP073989.D	21 Jul 2025 15:50		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP073990.D	21 Jul 2025 16:07		YPIAJ	Ok
21	I.BLK	I.BLK	PP073991.D	21 Jul 2025 16:23		YPIAJ	Ok
22	PB168927BL	PB168927BL	PP073992.D	21 Jul 2025 16:39		YPIAJ	Ok
23	PB168927BS	PB168927BS	PP073993.D	21 Jul 2025 16:56		YPIAJ	Ok
24	Q2638-01	OU4-TS-31-071725	PP073994.D	21 Jul 2025 17:12	DCB Low in both column	YPIAJ	ReRun
25	Q2638-03	OU4-TS-32-071725	PP073995.D	21 Jul 2025 17:29	DCB Low in both column	YPIAJ	ReRun
26	Q2638-05	OU4-TS-33-071725	PP073996.D	21 Jul 2025 17:45	DCB Low in 1st column	YPIAJ	Ok
27	Q2638-07	OU4-TS-34-071725	PP073997.D	21 Jul 2025 18:01	DCB Low in both column	YPIAJ	ReRun
28	Q2638-09	OU4-TS-35-071725	PP073998.D	21 Jul 2025 18:18		YPIAJ	Ok
29	Q2638-11	OU4-TS-36-071725	PP073999.D	21 Jul 2025 18:34	DCB Low in 1st column	YPIAJ	Ok
30	Q2638-13	OU4-TS-37-071725	PP074000.D	21 Jul 2025 18:50		YPIAJ	Ok
31	Q2639-01	OU4-TS-38-071725	PP074001.D	21 Jul 2025 19:07	DCB Low in 1st column,AR1254 hit	YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP074002.D	21 Jul 2025 20:12	AR1260-2 low in first column	YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP074003.D	21 Jul 2025 20:45		YPIAJ	Ok
34	AR1248CCC500	AR1248CCC500	PP074004.D	21 Jul 2025 21:01		YPIAJ	Ok,M
35	AR1254CCC500	AR1254CCC500	PP074005.D	21 Jul 2025 21:17		YPIAJ	Ok
36	I.BLK	I.BLK	PP074006.D	21 Jul 2025 21:34		YPIAJ	Ok
37	Q2663-05	HEI379F-1-1	PP074007.D	21 Jul 2025 21:50		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP072125

Review By	yogesh	Review On	7/21/2025 11:07:59 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:32:53 AM
SubDirectory	PP072125	HP Acquire Method	HP Processing Method PP070825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q2663-06	HEI379F-1-2	PP074008.D	21 Jul 2025 22:06		YPIAJ	Ok
39	Q2663-07	BC274318-1-1	PP074009.D	21 Jul 2025 22:23		YPIAJ	Ok,M
40	Q2663-08	BC274318-1-2	PP074010.D	21 Jul 2025 22:39	AR1260 hit	YPIAJ	Ok
41	Q2663-09	BC247637-1-1	PP074011.D	21 Jul 2025 22:55		YPIAJ	Ok,M
42	Q2663-10	BC247637-1-2	PP074012.D	21 Jul 2025 23:11		YPIAJ	Ok
43	Q2663-11	BC122903-1-1	PP074013.D	21 Jul 2025 23:28		YPIAJ	Ok,M
44	Q2663-12	BC122903-1-2	PP074014.D	21 Jul 2025 23:44		YPIAJ	Ok,M
45	Q2663-13	HEH700H-1-1	PP074015.D	22 Jul 2025 00:00	Signal#2 not detected.	YPIAJ	Not Ok
46	Q2663-14	HEH700H-1-2	PP074016.D	22 Jul 2025 00:17	AR1254 Hit	YPIAJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PP074017.D	22 Jul 2025 01:38		YPIAJ	Ok,M
48	AR1242CCC500	AR1242CCC500	PP074018.D	22 Jul 2025 01:55		YPIAJ	Ok,M
49	AR1248CCC500	AR1248CCC500	PP074019.D	22 Jul 2025 02:11		YPIAJ	Ok
50	AR1254CCC500	AR1254CCC500	PP074020.D	22 Jul 2025 02:27		YPIAJ	Ok
51	I.BLK	I.BLK	PP074021.D	22 Jul 2025 02:44		YPIAJ	Ok

M : Manual Integration