

Surrogate Summary

SDG No.: **Q2645**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112087.D	PIBLK-PO112087.D	Tetrachloro-m-xyl	1	20	19.4	97		60	140
		Decachlorobiphen	1	20	18.8	94		60	140
		Tetrachloro-m-xyl	2	20	20.0	100		60	140
		Decachlorobiphen	2	20	19.3	97		60	140
I.BLK-PO112336.D	PIBLK-PO112336.D	Tetrachloro-m-xyl	1	20	22.3	112		60	140
		Decachlorobiphen	1	20	18.4	92		60	140
		Tetrachloro-m-xyl	2	20	20.7	104		60	140
		Decachlorobiphen	2	20	17.7	88		60	140
Q2645-02	RW5B-CARBON-20250716	Tetrachloro-m-xyl	1	20	0.28	1	*	44	130
		Decachlorobiphen	1	20	1.12	6	*	60	125
		Tetrachloro-m-xyl	2	20	0.34	2	*	44	130
		Decachlorobiphen	2	20	1.10	6	*	60	125
I.BLK-PO112350.D	PIBLK-PO112350.D	Tetrachloro-m-xyl	1	20	20.1	101		60	140
		Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	19.0	95		60	140
		Decachlorobiphen	2	20	17.4	87		60	140
I.BLK-PO112370.D	PIBLK-PO112370.D	Tetrachloro-m-xyl	1	20	20.1	100		60	140
		Decachlorobiphen	1	20	17.7	89		60	140
		Tetrachloro-m-xyl	2	20	18.3	91		60	140
		Decachlorobiphen	2	20	17.4	87		60	140
PB168946BL	PB168946BL	Tetrachloro-m-xyl	1	20	20.9	104		44	130
		Decachlorobiphen	1	20	19.6	98		60	125
		Tetrachloro-m-xyl	2	20	18.2	91		44	130
		Decachlorobiphen	2	20	18.9	95		60	125
PB168946BS	PB168946BS	Tetrachloro-m-xyl	1	20	20.1	101		44	130
		Decachlorobiphen	1	20	20.3	102		60	125
		Tetrachloro-m-xyl	2	20	18.1	90		44	130
		Decachlorobiphen	2	20	19.5	98		60	125
Q2645-02RX	RW5B-CARBON-20250716RX	Tetrachloro-m-xyl	1	20	0.23	1	*	44	130
		Decachlorobiphen	1	20	1.09	5	*	60	125
		Tetrachloro-m-xyl	2	20	0.23	1	*	44	130
		Decachlorobiphen	2	20	0.26	1	*	60	125
Q2649-01MS	WC-1MS	Tetrachloro-m-xyl	1	20	24.5	122		44	130
		Decachlorobiphen	1	20	17.6	88		60	125
		Tetrachloro-m-xyl	2	20	22.2	111		44	130
		Decachlorobiphen	2	20	17.2	86		60	125
Q2649-01MSD	WC-1MSD	Tetrachloro-m-xyl	1	20	19.6	98		44	130
		Decachlorobiphen	1	20	17.8	89		60	125
		Tetrachloro-m-xyl	2	20	17.5	88		44	130
		Decachlorobiphen	2	20	17.4	87		60	125
I.BLK-PO112385.D	PIBLK-PO112385.D	Tetrachloro-m-xyl	1	20	19.8	99		60	140

Surrogate Summary

SDG No.: Q2645

Client: Tetra Tech NUS, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112385.D	PIBLK-PO112385.D	Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	18.1	90		60	140
I.BLK-PP073553.D	PIBLK-PP073553.D	Decachlorobiphen	2	20	17.4	87		60	140
		Tetrachloro-m-xyl	1	20	16.2	81		60	140
		Decachlorobiphen	1	20	17.3	87		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
I.BLK-PP073976.D	PIBLK-PP073976.D	Decachlorobiphen	2	20	17.1	85		60	140
		Tetrachloro-m-xyl	1	20	16.8	84		60	140
		Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	16.9	84		60	140
Q2635-01MS	RT2286MS	Decachlorobiphen	2	20	19.1	96		60	140
		Tetrachloro-m-xyl	1	20	16.8	84		44	130
		Decachlorobiphen	1	20	18.0	90		60	125
		Tetrachloro-m-xyl	2	20	17.9	89		44	130
Q2635-01MSD	RT2286MSD	Decachlorobiphen	2	20	19.9	99		60	125
		Tetrachloro-m-xyl	1	20	16.8	84		44	130
		Decachlorobiphen	1	20	17.8	89		60	125
		Tetrachloro-m-xyl	2	20	17.9	89		44	130
I.BLK-PP073991.D	PIBLK-PP073991.D	Decachlorobiphen	2	20	19.4	97		60	125
		Tetrachloro-m-xyl	1	20	15.8	79		60	140
		Decachlorobiphen	1	20	16.7	83		60	140
		Tetrachloro-m-xyl	2	20	17.5	87		60	140
PB168927BL	PB168927BL	Decachlorobiphen	2	20	19.7	98		60	140
		Tetrachloro-m-xyl	1	20	17.3	87		44	130
		Decachlorobiphen	1	20	18.7	94		60	125
		Tetrachloro-m-xyl	2	20	18.9	94		44	130
PB168927BS	PB168927BS	Decachlorobiphen	2	20	21.9	109		60	125
		Tetrachloro-m-xyl	1	20	18.8	94		44	130
		Decachlorobiphen	1	20	20.5	103		60	125
		Tetrachloro-m-xyl	2	20	19.3	96		44	130
I.BLK-PP074006.D	PIBLK-PP074006.D	Decachlorobiphen	2	20	23.9	120		60	125
		Tetrachloro-m-xyl	1	20	15.8	79		60	140
		Decachlorobiphen	1	20	17.1	85		60	140
		Tetrachloro-m-xyl	2	20	17.4	87		60	140
		Decachlorobiphen	2	20	19.3	97		60	140