

Surrogate Summary

SW-846

SDG No.: Q2645

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168948BL	PB168948BL	2-Fluorophenol	150	115	77		35	115
		Phenol-d6	150	116	77		34	127
		Nitrobenzene-d5	100	79.9	80		37	122
		2-Fluorobiphenyl	100	74.8	75		44	115
		2,4,6-Tribromophenol	150	125	83		39	132
		Terphenyl-d14	100	80.0	80		54	127
PB168948BS	PB168948BS	2-Fluorophenol	150	116	77		35	115
		Phenol-d6	150	118	79		34	127
		Nitrobenzene-d5	100	85.0	85		37	122
		2-Fluorobiphenyl	100	78.7	79		44	115
		2,4,6-Tribromophenol	150	131	87		39	132
		Terphenyl-d14	100	85.7	86		54	127
Q2645-02	RW5B-CARBON-20250716	2-Fluorophenol	150	3.73	2	*	35	115
		Phenol-d6	150	3.88	3	*	34	127
		Nitrobenzene-d5	100	9.31	9	*	37	122
		2-Fluorobiphenyl	100	6.94	7	*	44	115
		2,4,6-Tribromophenol	150	4.54	3	*	39	132
		Terphenyl-d14	100	4.04	4	*	54	127
Q2649-09MS	WC-3MS	2-Fluorophenol	150	81.7	54		35	115
		Phenol-d6	150	89.3	60		34	127
		Nitrobenzene-d5	100	58.8	59		37	122
		2-Fluorobiphenyl	100	54.3	54		44	115
		2,4,6-Tribromophenol	150	92.4	62		39	132
		Terphenyl-d14	100	56.4	56		54	127
Q2649-09MSD	WC-3MSD	2-Fluorophenol	150	83.8	56		35	115
		Phenol-d6	150	92.0	61		34	127
		Nitrobenzene-d5	100	63.6	64		37	122
		2-Fluorobiphenyl	100	54.0	54		44	115
		2,4,6-Tribromophenol	150	92.4	62		39	132
		Terphenyl-d14	100	57.8	58		54	127