

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143138.D
Instrument ID: BNA_F

Contract: TETR06
SDG NO.: Q2645
DFTPP Injection Date: 07/17/2025
DFTPP Injection Time: 09:58

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.9) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	79.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF143140.D	07/17/2025	11:04
SSTDICC005	SSTDICC005	BF143141.D	07/17/2025	11:34
SSTDICC010	SSTDICC010	BF143142.D	07/17/2025	12:04
SSTDICC020	SSTDICC020	BF143143.D	07/17/2025	12:34
SSTDICCC040	SSTDICCC040	BF143144.D	07/17/2025	13:03
SSTDICC050	SSTDICC050	BF143145.D	07/17/2025	13:33
SSTDICC060	SSTDICC060	BF143146.D	07/17/2025	14:04
SSTDICC080	SSTDICC080	BF143147.D	07/17/2025	14:34