

Surrogate Summary

SDG No.: Q2645

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2645-03	RW5B-CARBON-20250716	1,2-Dichloroethane-d4	50	56.3	113		81	118
		Dibromofluoromethane	50	52.0	104		80	119
		Toluene-d8	50	51.0	102		89	112
		4-Bromofluorobenzene	50	46.2	92		85	114
VN0721WBL01	VN0721WBL01	1,2-Dichloroethane-d4	50	54.7	109		81	118
		Dibromofluoromethane	50	51.4	103		80	119
		Toluene-d8	50	51.0	102		89	112
		4-Bromofluorobenzene	50	46.4	93		85	114
VN0721WBS01	VN0721WBS01	1,2-Dichloroethane-d4	50	47.3	95		81	118
		Dibromofluoromethane	50	48.8	98		80	119
		Toluene-d8	50	50.4	101		89	112
		4-Bromofluorobenzene	50	49.9	100		85	114
VN0721WBSD0	VN0721WBSD01	1,2-Dichloroethane-d4	50	47.2	94		81	118
		Dibromofluoromethane	50	50.3	101		80	119
		Toluene-d8	50	50.6	101		89	112
		4-Bromofluorobenzene	50	50.0	100		85	114