

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2645</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VN087371.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0721WBS01	Vinyl chloride	20	19.0	ug/L	95			58	137	
	1,1-Dichloroethene	20	18.1	ug/L	91			71	131	
	2-Butanone	100	86.4	ug/L	86			56	143	
	Carbon Tetrachloride	20	19.6	ug/L	98			72	136	
	Chloroform	20	17.9	ug/L	90			79	124	
	Benzene	20	19.4	ug/L	97			79	120	
	1,2-Dichloroethane	20	18.3	ug/L	92			73	128	
	Trichloroethene	20	18.8	ug/L	94			79	123	
	Tetrachloroethene	20	18.3	ug/L	92			74	129	
	Chlorobenzene	20	18.6	ug/L	93			82	118	