

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2645</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VN087372.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0721WBSD01	Vinyl chloride	20	19.0	ug/L	95	0		58	137	20
	1,1-Dichloroethene	20	17.6	ug/L	88	3		71	131	20
	2-Butanone	100	88.2	ug/L	88	2		56	143	20
	Carbon Tetrachloride	20	19.8	ug/L	99	1		72	136	20
	Chloroform	20	18.4	ug/L	92	2		79	124	20
	Benzene	20	19.2	ug/L	96	1		79	120	20
	1,2-Dichloroethane	20	18.8	ug/L	94	2		73	128	20
	Trichloroethene	20	18.8	ug/L	94	0		79	123	20
	Tetrachloroethene	20	19.3	ug/L	97	5		74	129	20
	Chlorobenzene	20	18.6	ug/L	93	0		82	118	20