

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Lab File ID: VN087327.D

BFB Injection Date: 07/16/2025

Instrument ID: MSVOA_N

BFB Injection Time: 16:10

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.8
75	30.0 - 60.0% of mass 95	50.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.1
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	70.9
175	5.0 - 9.0% of mass 174	3.6 (5.1) 1
176	95.0 - 101.0% of mass 174	68.7 (96.9) 1
177	5.0 - 9.0% of mass 176	4.8 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN087328.D	07/16/2025	17:05
VSTDIC005	VSTDIC005	VN087329.D	07/16/2025	17:27
VSTDIC020	VSTDIC020	VN087330.D	07/16/2025	17:49
VSTDIC050	VSTDIC050	VN087331.D	07/16/2025	18:11
VSTDIC100	VSTDIC100	VN087332.D	07/16/2025	18:32
VSTDIC150	VSTDIC150	VN087333.D	07/16/2025	18:54