

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q2645

Lab File ID: VN087367.D

BFB Injection Date: 07/21/2025

Instrument ID: MSVOA_N

BFB Injection Time: 10:05

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.1
75	30.0 - 60.0% of mass 95	52.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.4
173	Less than 2.0% of mass 174	1.5 (1.9) 1
174	50.0 - 100.0% of mass 95	75.2
175	5.0 - 9.0% of mass 174	6 (8) 1
176	95.0 - 101.0% of mass 174	72.4 (96.3) 1
177	5.0 - 9.0% of mass 176	5 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN087368.D	07/21/2025	10:38
VN0721WBL01	VN0721WBL01	VN087370.D	07/21/2025	11:34
VN0721WBS01	VN0721WBS01	VN087371.D	07/21/2025	11:55
VN0721WBSD01	VN0721WBSD01	VN087372.D	07/21/2025	12:30
RW5B-CARBON-20250716	Q2645-03	VN087377.D	07/21/2025	14:16
VSTDCCC050EC	VSTDCCC050	VN087387.D	07/21/2025	17:50