

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136549

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/21/2025	09:05
2	CAL2	1	Water	NA	NA	20.2	7.00	07/21/2025	09:06
3	CAL3	1	Water	NA	NA	20.3	10.02	07/21/2025	09:10
4	ICV	1	Water	NA	NA	20.2	7.01	07/21/2025	09:11
5	CCV1	1	Water	NA	NA	20.2	2.01	07/21/2025	09:15
6	Q2640-01	1	Water	NA	NA	22.1	8.83	07/21/2025	09:30
7	Q2640-01DUP	1	Water	NA	NA	22.2	8.84	07/21/2025	09:33
8	Q2646-01	1	Water	NA	NA	20.7	4.66	07/21/2025	09:44
9	CCV2	1	Water	NA	NA	20.2	12.02	07/21/2025	09:45

WORKLIST(Hardcopy Internal Chain)

136549

WorkList Name : PH Q2640 WorkList ID : 190841 Department : Wet-Chemistry Date : 07-21-2025 08:46:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2640-01	ELMORA-WATER	Water	pH	Cool 4 deg C	PSEG03	D41	07/18/2025	9040C
Q2646-01	FRAC TANK	Water	pH	Cool 4 deg C	ENV160	O11	07/18/2025	9040C

Date/Time 07/21/25 09:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 07/21/25 13:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]