

Duplicate Sample Summary

Client: Environmental Restoration, LLC Project: North Arlington, NJ Client ID: ELMORA-WATERDUP	SDG No.: Q2646 Sample ID: Q2640-01 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
pH	pH	+/-20	8.83		8.84		1	0.11		07/21/2025
Flash Point	o F	+/-2	>212.0		>212.0		1	0		07/21/2025

Duplicate Sample Summary

Client:	Environmental Restoration, LLC	SDG No.:	Q2646
Project:	North Arlington, NJ	Sample ID:	Q2646-01
Client ID:	FRAC TANKDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	562		559		1	0.54		07/23/2025