

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2646

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168961BL		WATER		Batch Number:	PB168961		Prep Date:	07/21/2025	
	Mercury	0.076	<0.2	U	0.20	CV	07/22/2025	17:21	LB136570

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2646

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168932BL	WATER			Batch Number:	PB168932		Prep Date:	07/21/2025	
	Aluminum	5.67	<25	U	50.0	P	07/22/2025	15:05	LB136571
	Antimony	3.38	<12.5	U	25.0	P	07/22/2025	15:05	LB136571
	Arsenic	2.56	<5	U	10.0	P	07/22/2025	15:05	LB136571
	Barium	7.28	<25	U	50.0	P	07/22/2025	15:05	LB136571
	Beryllium	0.28	<1.5	U	3.00	P	07/22/2025	15:05	LB136571
	Cadmium	0.25	<1.5	U	3.00	P	07/22/2025	15:05	LB136571
	Calcium	117	<500	U	1000	P	07/22/2025	15:05	LB136571
	Chromium	1.06	<2.5	U	5.00	P	07/22/2025	15:05	LB136571
	Cobalt	1.13	<7.5	U	15.0	P	07/22/2025	15:05	LB136571
	Copper	2.30	<5	U	10.0	P	07/22/2025	15:05	LB136571
	Iron	11.7	<25	U	50.0	P	07/22/2025	15:05	LB136571
	Lead	1.15	<3	U	6.00	P	07/22/2025	15:05	LB136571
	Magnesium	122	<500	U	1000	P	07/22/2025	15:05	LB136571
	Manganese	2.97	<5	U	10.0	P	07/22/2025	15:05	LB136571
	Nickel	1.53	<10	U	20.0	P	07/22/2025	15:05	LB136571
	Potassium	459	<500	U	1000	P	07/22/2025	15:05	LB136571
	Selenium	4.82	<5	U	10.0	P	07/22/2025	15:05	LB136571
	Silver	0.81	<2.5	U	5.00	P	07/22/2025	15:05	LB136571
	Sodium	434	<500	U	1000	P	07/22/2025	15:05	LB136571
	Thallium	2.19	<10	U	20.0	P	07/22/2025	15:05	LB136571
	Vanadium	3.13	<10	U	20.0	P	07/22/2025	15:05	LB136571
	Zinc	8.33	<10	U	20.0	P	07/22/2025	15:05	LB136571