

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2646

Contract: ENVI60

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168932BS							
Aluminum	ug/L	1000	1050		105	86 - 115	P
Antimony	ug/L	400	431		108	88 - 113	P
Arsenic	ug/L	400	403		101	87 - 113	P
Barium	ug/L	100	108		108	88 - 113	P
Beryllium	ug/L	100	106		106	89 - 112	P
Cadmium	ug/L	100	97.7		98	88 - 113	P
Calcium	ug/L	500	559	J	112	87 - 113	P
Chromium	ug/L	200	208		104	90 - 113	P
Cobalt	ug/L	100	103		103	89 - 114	P
Copper	ug/L	150	162		108	86 - 114	P
Iron	ug/L	1500	1580		105	87 - 115	P
Lead	ug/L	500	491		98	86 - 113	P
Magnesium	ug/L	1000	1060		106	85 - 113	P
Manganese	ug/L	100	109		109	90 - 114	P
Nickel	ug/L	250	260		104	88 - 113	P
Potassium	ug/L	5000	5170		103	86 - 114	P
Selenium	ug/L	1000	1020		102	83 - 114	P
Silver	ug/L	37.5	40.0		107	84 - 115	P
Sodium	ug/L	1500	1480		99	87 - 115	P
Thallium	ug/L	1000	1070		107	85 - 114	P
Vanadium	ug/L	150	157		105	90 - 111	P
Zinc	ug/L	100	110		110	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Environmental Restoration, LLC

SDG No.: Q2646

Contract: ENVI60

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168961BS Mercury	ug/L	4.0	3.84		96	80 - 120	CV