



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Environmental Restoration, LLC

SDG No.: Q2646

Contract: ENVI60

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	20.5	20.0	103	65 - 135	P	07/22/2025	12:32	LB136571
	Barium	99.5	100	100	65 - 135	P	07/22/2025	12:32	LB136571
	Cadmium	5.65	6.0	94	65 - 135	P	07/22/2025	12:32	LB136571
	Chromium	10.5	10.0	106	65 - 135	P	07/22/2025	12:32	LB136571
	Lead	12.4	12.0	103	65 - 135	P	07/22/2025	12:32	LB136571
	Selenium	21.2	20.0	106	65 - 135	P	07/22/2025	12:32	LB136571
	Silver	9.69	10.0	97	65 - 135	P	07/22/2025	12:32	LB136571
CRA	Mercury	0.19	0.2	93	70 - 130	CV	07/22/2025	15:22	LB136570