

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112396.D **Time Analyzed:** 22:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.757	4.661	4.861	502.730	500.000	0.5
Aroclor-1016-2	4.776	4.680	4.880	503.830	500.000	0.8
Aroclor-1016-3	4.832	4.736	4.936	492.930	500.000	-1.4
Aroclor-1016-4	4.952	4.856	5.056	497.900	500.000	-0.4
Aroclor-1016-5	5.209	5.113	5.313	543.660	500.000	8.7
Aroclor-1260-1	6.246	6.151	6.351	475.920	500.000	-4.8
Aroclor-1260-2	6.436	6.341	6.541	482.330	500.000	-3.5
Aroclor-1260-3	6.802	6.707	6.907	482.820	500.000	-3.4
Aroclor-1260-4	7.061	6.967	7.167	499.400	500.000	-0.1
Aroclor-1260-5	7.304	7.209	7.409	492.720	500.000	-1.5
Decachlorobiphenyl	8.695	8.603	8.803	49.860	50.000	-0.3
Tetrachloro-m-xylene	3.670	3.573	3.773	54.520	50.000	9.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112396.D **Time Analyzed:** 22:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.661	4.861	430.080	500.000	-14.0
Aroclor-1016-2	4.756	4.679	4.879	458.520	500.000	-8.3
Aroclor-1016-3	4.931	4.854	5.054	457.570	500.000	-8.5
Aroclor-1016-4	4.974	4.897	5.097	451.460	500.000	-9.7
Aroclor-1016-5	5.186	5.109	5.309	500.960	500.000	0.2
Aroclor-1260-1	6.215	6.138	6.338	467.980	500.000	-6.4
Aroclor-1260-2	6.403	6.326	6.526	489.800	500.000	-2.0
Aroclor-1260-3	6.554	6.478	6.678	483.920	500.000	-3.2
Aroclor-1260-4	7.024	6.948	7.148	454.700	500.000	-9.1
Aroclor-1260-5	7.266	7.190	7.390	472.510	500.000	-5.5
Decachlorobiphenyl	8.642	8.568	8.768	46.870	50.000	-6.3
Tetrachloro-m-xylene	3.663	3.584	3.784	49.510	50.000	-1.0