

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112410.D **Time Analyzed:** 04:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.753	4.661	4.861	510.700	500.000	2.1
Aroclor-1016-2	4.772	4.680	4.880	515.460	500.000	3.1
Aroclor-1016-3	4.829	4.736	4.936	513.700	500.000	2.7
Aroclor-1016-4	4.949	4.856	5.056	527.620	500.000	5.5
Aroclor-1016-5	5.206	5.113	5.313	433.090	500.000	-13.4
Aroclor-1260-1	6.243	6.151	6.351	437.650	500.000	-12.5
Aroclor-1260-2	6.432	6.341	6.541	464.600	500.000	-7.1
Aroclor-1260-3	6.798	6.707	6.907	415.110	500.000	-17.0
Aroclor-1260-4	7.058	6.967	7.167	441.550	500.000	-11.7
Aroclor-1260-5	7.300	7.209	7.409	415.340	500.000	-16.9
Decachlorobiphenyl	8.692	8.603	8.803	38.770	50.000	-22.5
Tetrachloro-m-xylene	3.668	3.573	3.773	54.490	50.000	9.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112410.D **Time Analyzed:** 04:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.736	4.661	4.861	466.220	500.000	-6.8
Aroclor-1016-2	4.754	4.679	4.879	467.910	500.000	-6.4
Aroclor-1016-3	4.929	4.854	5.054	481.890	500.000	-3.6
Aroclor-1016-4	4.974	4.897	5.097	357.030	500.000	-28.6
Aroclor-1016-5	5.184	5.109	5.309	447.580	500.000	-10.5
Aroclor-1260-1	6.212	6.138	6.338	438.580	500.000	-12.3
Aroclor-1260-2	6.400	6.326	6.526	485.090	500.000	-3.0
Aroclor-1260-3	6.552	6.478	6.678	448.230	500.000	-10.4
Aroclor-1260-4	7.022	6.948	7.148	406.170	500.000	-18.8
Aroclor-1260-5	7.264	7.190	7.390	393.210	500.000	-21.4
Decachlorobiphenyl	8.639	8.568	8.768	39.000	50.000	-22.0
Tetrachloro-m-xylene	3.662	3.584	3.784	50.260	50.000	0.5