



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2646

Analytical Method: 8082A

Client: Environmental Restoration, LLC

Datafile : PO112403.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168957BSD (Column 1)	AR1016	5	5.10	ug/L	102	6			40 (77)	140 (107)	20 (20)
	AR1260	5	4.40	ug/L	88	5			40 (66)	140 (113)	20 (20)
PB168957BSD (Column 2)	AR1016	5	4.70	ug/L	94	7			40 (77)	140 (107)	20 (20)
	AR1260	5	4.50	ug/L	90	7			40 (66)	140 (113)	20 (20)