

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG NO.: Q2646

Continuing Calib Date: 07/24/2025

Initial Calibration Date(s): 07/21/2025 07/21/2025

Continuing Calib Time: 23:29

Initial Calibration Time(s): 15:02 16:39

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.32	7.33	7.23	7.43	0.01
2,4-D	8.45	8.46	8.36	8.56	0.01
2,4,5-TP(Silvex)	9.34	9.34	9.24	9.44	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG NO.: Q2646

Continuing Calib Date: 07/24/2025

Initial Calibration Date(s): 07/21/2025 07/21/2025

Continuing Calib Time: 23:29

Initial Calibration Time(s): 15:02 16:39

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
2,4-DCAA	7.76	7.77	7.67	7.87	0.01
2,4-D	9.02	9.02	8.92	9.12	0.00
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00