

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG NO.: Q2646

Continuing Calib Date: 07/31/2025

Initial Calibration Date(s): 07/29/2025 07/29/2025

Continuing Calib Time: 15:43

Initial Calibration Time(s): 16:30 18:07

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.31	7.32	7.22	7.42	0.02
2,4-D	8.43	8.45	8.35	8.55	0.02
2,4,5-TP(Silvex)	9.31	9.33	9.23	9.43	0.02

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG NO.: Q2646

Continuing Calib Date: 07/31/2025

Initial Calibration Date(s): 07/29/2025 07/29/2025

Continuing Calib Time: 15:43

Initial Calibration Time(s): 16:30 18:07

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
2,4-DCAA	7.76	7.77	7.67	7.87	0.01
2,4-D	9.01	9.03	8.93	9.13	0.02
2,4,5-TP(Silvex)	9.91	9.92	9.82	10.02	0.01