

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ENVI60  
**Lab Code:** ACE **SDG NO.:** Q2646  
**GC Column:** RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 07/31/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS031336.D **Time Analyzed:** 12:00

| COMPOUND         | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|------------------|-------|----------------------|-------|--------------------|-------------------|-----|
| 2,4,5-TP(Silvex) | 9.315 | 9.227                | 9.427 | 745.440            | 712.500           | 4.6 |
| 2,4-D            | 8.436 | 8.346                | 8.546 | 738.480            | 705.000           | 4.7 |
| 2,4-DCAA         | 7.306 | 7.215                | 7.415 | 766.150            | 750.000           | 2.2 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ENVI60  
**Lab Code:** ACE **SDG NO.:** Q2646  
**GC Column:** RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 07/31/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS031336.D **Time Analyzed:** 12:00

| COMPOUND         | RT    | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|-------|-----------|--------|--------------------|-------------------|------|
|                  |       | FROM      | TO     |                    |                   |      |
| 2,4,5-TP(Silvex) | 9.916 | 9.824     | 10.024 | 692.060            | 712.500           | -2.9 |
| 2,4-D            | 9.018 | 8.925     | 9.125  | 670.620            | 705.000           | -4.9 |
| 2,4-DCAA         | 7.760 | 7.666     | 7.866  | 708.740            | 750.000           | -5.5 |