

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/31/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031339.D **Time Analyzed:** 15:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.314	9.227	9.427	793.360	712.500	11.3
2,4-D	8.433	8.346	8.546	815.510	705.000	15.7
2,4-DCAA	7.305	7.215	7.415	810.300	750.000	8.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/31/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031339.D **Time Analyzed:** 15:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.914	9.824	10.024	708.190	712.500	-0.6
2,4-D	9.014	8.925	9.125	717.030	705.000	1.7
2,4-DCAA	7.756	7.666	7.866	757.070	750.000	0.9