

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2646</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Environmental Restoration, LLC</u>	DataFile :	<u>PS031234.D</u>

		Sample				Rec		RPD		Limits	
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High
Lab Sample ID:	Q2641-02MS	Client Sample ID:		P001-CONCRETE001-01M							
	(Column 1)										
	2,4-D	50	0	80.3	ug/L	161	*			70 (65)	130 (135)
	2,4,5-TP(Silvex)	50	0	77.7	ug/L	155	*			70 (62)	130 (139)
Lab Sample ID:	Q2641-02MS	Client Sample ID:		P001-CONCRETE001-01M							
	(Column 2)										
	2,4-D	50	0	86.2	ug/L	172	*			70 (65)	130 (135)
	2,4,5-TP(Silvex)	50	0	76.7	ug/L	153	*			70 (62)	130 (139)