

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2646</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Environmental Restoration, LLC</u>	DataFile :	<u>PS031235.D</u>

		Sample					Rec		RPD		Limits	
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2641-02MSD (Column 1)	Client Sample ID:	P001-CONCRETE001-01M									
	2,4-D	50	0	76.2	ug/L	152	*	6		70 (65)	130 (135)	20 (20)
	2,4,5-TP(Silvex)	50	0	73.8	ug/L	148	*	5		70 (62)	130 (139)	20 (20)
Lab Sample ID:	Q2641-02MSD (Column 2)	Client Sample ID:	P001-CONCRETE001-01M									
	2,4-D	50	0	80.9	ug/L	162	*	6		70 (65)	130 (135)	20 (20)
	2,4,5-TP(Silvex)	50	0	72.7	ug/L	145	*	5		70 (62)	130 (139)	20 (20)