

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/24/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096553.D **Time Analyzed:** 11:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.017	8.920	9.120	54.750	50.000	9.5
Endrin	6.543	6.445	6.645	52.690	50.000	5.4
gamma-BHC (Lindane)	4.311	4.213	4.413	57.100	50.000	14.2
Heptachlor	4.903	4.806	5.006	54.190	50.000	8.4
Heptachlor epoxide	5.662	5.565	5.765	57.860	50.000	15.7
Methoxychlor	7.463	7.364	7.564	57.890	50.000	15.8
Tetrachloro-m-xylene	3.535	3.438	3.638	55.150	50.000	10.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/24/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096553.D **Time Analyzed:** 11:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.993	7.896	8.096	47.090	50.000	-5.8
Endrin	5.709	5.612	5.812	56.050	50.000	12.1
gamma-BHC (Lindane)	3.664	3.567	3.767	57.790	50.000	15.6
Heptachlor	4.013	3.917	4.117	55.010	50.000	10.0
Heptachlor epoxide	4.798	4.702	4.902	56.250	50.000	12.5
Methoxychlor	6.681	6.584	6.784	54.670	50.000	9.3
Tetrachloro-m-xylene	2.827	2.729	2.929	57.210	50.000	14.4