

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096667.D **Time Analyzed:** 14:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.015	8.916	9.116	55.130	50.000	10.3
Endrin	6.542	6.442	6.642	52.120	50.000	4.2
gamma-BHC (Lindane)	4.310	4.210	4.410	53.830	50.000	7.7
Heptachlor	4.902	4.803	5.003	54.410	50.000	8.8
Heptachlor epoxide	5.662	5.562	5.762	55.110	50.000	10.2
Methoxychlor	7.461	7.362	7.562	50.160	50.000	0.3
Tetrachloro-m-xylene	3.535	3.435	3.635	53.970	50.000	7.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096667.D **Time Analyzed:** 14:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.992	7.892	8.092	58.200	50.000	16.4
Endrin	5.709	5.609	5.809	52.210	50.000	4.4
gamma-BHC (Lindane)	3.665	3.565	3.765	54.890	50.000	9.8
Heptachlor	4.014	3.914	4.114	53.610	50.000	7.2
Heptachlor epoxide	4.799	4.699	4.899	54.860	50.000	9.7
Methoxychlor	6.681	6.581	6.781	51.330	50.000	2.7
Tetrachloro-m-xylene	2.827	2.728	2.928	54.460	50.000	8.9