

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ENVI60
Lab Code: ACE
SDG NO.: Q2646
GC Column: ZB-MR1
ID: 0.32 (mm)
 Initi. Calib. Date(s): 07/28/2025 07/28/2025

Client Sample No.: CCAL04
Date Analyzed: 07/31/2025
Lab Sample No.: PSTDCCC050
Data File : PL096678.D
Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.014	8.916	9.116	56.240	50.000	12.5
Endrin	6.542	6.442	6.642	50.770	50.000	1.5
gamma-BHC (Lindane)	4.310	4.210	4.410	53.420	50.000	6.8
Heptachlor	4.902	4.803	5.003	53.720	50.000	7.4
Heptachlor epoxide	5.661	5.562	5.762	54.510	50.000	9.0
Methoxychlor	7.461	7.362	7.562	48.560	50.000	-2.9
Tetrachloro-m-xylene	3.534	3.435	3.635	53.530	50.000	7.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2646
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/28/2025 07/28/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/31/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096678.D **Time Analyzed:** 17:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.992	7.892	8.092	52.490	50.000	5.0
Endrin	5.709	5.609	5.809	49.600	50.000	-0.8
gamma-BHC (Lindane)	3.665	3.565	3.765	53.920	50.000	7.8
Heptachlor	4.014	3.914	4.114	51.970	50.000	3.9
Heptachlor epoxide	4.799	4.699	4.899	53.430	50.000	6.9
Methoxychlor	6.680	6.581	6.781	46.140	50.000	-7.7
Tetrachloro-m-xylene	2.828	2.728	2.928	53.740	50.000	7.5