

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q2646

**Analytical Method:** 8081B

**Client:** Environmental Restoration, LLC

**Datafile :** PL096555.D

| Lab Sample ID            | Parameter           | Spike | Result | Units | Rec | RPD | Qual | RPD  | Limits  |           | RPD |
|--------------------------|---------------------|-------|--------|-------|-----|-----|------|------|---------|-----------|-----|
|                          |                     |       |        |       |     |     |      | Qual | Low     | High      |     |
| PB168984BS<br>(Column 1) | gamma-BHC (Lindane) | 0.5   | 0.56   | ug/L  | 113 |     |      |      | 40 (82) | 140 (129) |     |
|                          | Heptachlor          | 0.5   | 0.53   | ug/L  | 107 |     |      |      | 40 (79) | 140 (127) |     |
|                          | Heptachlor epoxide  | 0.5   | 0.61   | ug/L  | 122 |     |      |      | 40 (81) | 140 (124) |     |
|                          | Endrin              | 0.5   | 0.46   | ug/L  | 93  |     |      |      | 40 (81) | 140 (128) |     |
|                          | Methoxychlor        | 0.5   | 0.40   | ug/L  | 79  |     |      |      | 40 (78) | 140 (108) |     |
| PB168984BS<br>(Column 2) | gamma-BHC (Lindane) | 0.5   | 0.60   | ug/L  | 120 |     |      |      | 40 (82) | 140 (129) |     |
|                          | Heptachlor          | 0.5   | 0.55   | ug/L  | 110 |     |      |      | 40 (79) | 140 (127) |     |
|                          | Heptachlor epoxide  | 0.5   | 0.59   | ug/L  | 118 |     |      |      | 40 (81) | 140 (124) |     |
|                          | Endrin              | 0.5   | 0.48   | ug/L  | 95  |     |      |      | 40 (81) | 140 (128) |     |
|                          | Methoxychlor        | 0.5   | 0.40   | ug/L  | 79  |     |      |      | 40 (78) | 140 (108) |     |