

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071725\
 Data File : BF143143.D
 Acq On : 17 Jul 2025 12:34
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020

Quant Time: Jul 17 15:02:41 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 17 14:54:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.963	152	168491	20.000	ng	0.00
21) Naphthalene-d8	8.245	136	642580	20.000	ng	0.00
39) Acenaphthene-d10	9.998	164	329670	20.000	ng	0.00
64) Phenanthrene-d10	11.480	188	532921	20.000	ng	0.00
76) Chrysene-d12	14.121	240	305398	20.000	ng	0.00
86) Perylene-d12	15.621	264	338518	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.581	112	444546	41.752	ng	0.00
7) Phenol-d6	6.575	99	558427	41.668	ng	-0.01
23) Nitrobenzene-d5	7.516	82	554482	43.013	ng	0.00
42) 2,4,6-Tribromophenol	10.780	330	139748	41.034	ng	0.00
45) 2-Fluorobiphenyl	9.316	172	1019688	42.301	ng	0.00
79) Terphenyl-d14	13.063	244	866830	42.238	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.810	88	103481	20.540	ng	99
3) Pyridine	3.581	79	261405	19.990	ng	99
4) n-Nitrosodimethylamine	3.510	42	134638	19.942	ng	99
6) Aniline	6.622	93	382991	20.505	ng	99
8) 2-Chlorophenol	6.745	128	231044	20.702	ng	98
9) Benzaldehyde	6.510	77	196003	19.504	ng	98
10) Phenol	6.593	94	322193	22.068	ng	97
11) bis(2-Chloroethyl)ether	6.692	93	232425	20.792	ng	98
12) 1,3-Dichlorobenzene	6.904	146	251734	20.763	ng	99
13) 1,4-Dichlorobenzene	6.981	146	253137	20.733	ng	100
14) 1,2-Dichlorobenzene	7.134	146	243437	20.963	ng	99
15) Benzyl Alcohol	7.092	79	210016	20.524	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.234	45	428827	20.680	ng	100
17) 2-Methylphenol	7.198	107	192172	20.479	ng	99
18) Hexachloroethane	7.481	117	88036	20.828	ng	98
19) n-Nitroso-di-n-propyla...	7.363	70	173142	20.137	ng	97
20) 3+4-Methylphenols	7.351	107	247578	21.745	ng	96
22) Acetophenone	7.363	105	321774	21.151	ng	# 99
24) Nitrobenzene	7.534	77	249251	20.739	ng	98
25) Isophorone	7.775	82	457146	20.088	ng	99
26) 2-Nitrophenol	7.851	139	105681	20.261	ng	98
27) 2,4-Dimethylphenol	7.887	122	218960	20.594	ng	99
28) bis(2-Chloroethoxy)met...	7.987	93	283946	20.811	ng	100
29) 2,4-Dichlorophenol	8.092	162	184432	20.569	ng	99
30) 1,2,4-Trichlorobenzene	8.181	180	204103	21.070	ng	98
31) Naphthalene	8.263	128	662968	20.949	ng	100
32) Benzoic acid	7.957	122	90732	18.667	ng	98
33) 4-Chloroaniline	8.304	127	269101	20.825	ng	99
34) Hexachlorobutadiene	8.386	225	122624	20.722	ng	98
35) Caprolactam	8.651	113	53950	19.911	ng	95
36) 4-Chloro-3-methylphenol	8.775	107	197317	20.246	ng	99
37) 2-Methylnaphthalene	8.957	142	403298	20.943	ng	100
38) 1-Methylnaphthalene	9.057	142	417629	21.060	ng	100
40) 1,2,4,5-Tetrachloroben...	9.122	216	198009	20.804	ng	99
41) Hexachlorocyclopentadiene	9.116	237	122922	19.848	ng	100
43) 2,4,6-Trichlorophenol	9.228	196	135703	20.601	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF071725\
 Data File : BF143143.D
 Acq On : 17 Jul 2025 12:34
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020

Quant Time: Jul 17 15:02:41 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 17 14:54:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.263	196	135634	20.584	ng	99
46) 1,1'-Biphenyl	9.416	154	540637	21.019	ng	100
47) 2-Chloronaphthalene	9.445	162	393002	20.650	ng	99
48) 2-Nitroaniline	9.528	65	122517	20.528	ng	99
49) Acenaphthylene	9.857	152	663048	20.789	ng	100
50) Dimethylphthalate	9.710	163	441768	20.558	ng	100
51) 2,6-Dinitrotoluene	9.769	165	89742	20.529	ng	97
52) Acenaphthene	10.033	154	390249	20.702	ng	99
53) 3-Nitroaniline	9.933	138	103870	20.314	ng	94
54) 2,4-Dinitrophenol	10.039	184	26820	19.073	ng	# 1
55) Dibenzofuran	10.204	168	582320	20.806	ng	98
56) 4-Nitrophenol	10.081	139	75275	19.715	ng	99
57) 2,4-Dinitrotoluene	10.169	165	112851	20.577	ng	99
58) Fluorene	10.545	166	435866	20.750	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.316	232	113851	20.859	ng	100
60) Diethylphthalate	10.410	149	440058	20.719	ng	100
61) 4-Chlorophenyl-phenyle...	10.539	204	216496	20.699	ng	99
62) 4-Nitroaniline	10.545	138	87666	20.005	ng	99
63) Azobenzene	10.698	77	458013	20.690	ng	99
65) 4,6-Dinitro-2-methylph...	10.580	198	46122	20.060	ng	95
66) n-Nitrosodiphenylamine	10.651	169	384467	20.488	ng	100
67) 4-Bromophenyl-phenylether	11.028	248	128663	20.259	ng	99
68) Hexachlorobenzene	11.098	284	134031	20.191	ng	99
69) Atrazine	11.169	200	105386	20.370	ng	99
70) Pentachlorophenol	11.280	266	77906	19.957	ng	99
71) Phenanthrene	11.510	178	595411	21.042	ng	99
72) Anthracene	11.557	178	604416	20.944	ng	99
73) Carbazole	11.704	167	523365	20.769	ng	100
74) Di-n-butylphthalate	12.039	149	601867	21.108	ng	100
75) Fluoranthene	12.692	202	542662	20.894	ng	99
77) Benzidine	12.810	184	257904	21.919	ng	99
78) Pyrene	12.922	202	544650	20.897	ng	99
80) Butylbenzylphthalate	13.539	149	163546	20.491	ng	99
81) Benzo(a)anthracene	14.110	228	421791	20.629	ng	99
82) 3,3'-Dichlorobenzidine	14.068	252	151153	22.181	ng	100
83) Chrysene	14.145	228	380799	20.714	ng	99
84) Bis(2-ethylhexyl)phtha...	14.098	149	251739	20.839	ng	99
85) Di-n-octyl phthalate	14.716	149	436102	19.916	ng	99
87) Indeno(1,2,3-cd)pyrene	17.151	276	484240	20.286	ng	99
88) Benzo(b)fluoranthene	15.174	252	435747	21.071	ng	99
89) Benzo(k)fluoranthene	15.204	252	369899	20.044	ng	100
90) Benzo(a)pyrene	15.557	252	388136	20.371	ng	99
91) Dibenzo(a,h)anthracene	17.168	278	398949	20.534	ng	99
92) Benzo(g,h,i)perylene	17.615	276	385915	20.534	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 17 15:02:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 17 14:54:43 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.963 min Scan# 811
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:152 Resp: 168491
Ion Ratio Lower Upper
152 100
150 157.5 128.2 192.4
115 60.1 47.9 71.9



#2
1,4-Dioxane
Concen: 20.540 ng
RT: 2.810 min Scan# 105
Delta R.T. -0.012 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion: 88 Resp: 103481
Ion Ratio Lower Upper
88 100
58 86.7 70.2 105.2
43 33.8 27.1 40.7





#3

Pyridine

Concen: 19.990 ng

RT: 3.581 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion: 79 Resp: 261405

Ion Ratio Lower Upper

79 100

52 85.9 69.4 104.2

51 42.1 33.5 50.3



#4

n-Nitrosodimethylamine

Concen: 19.942 ng

RT: 3.510 min Scan# 224

Delta R.T. -0.023 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion: 42 Resp: 134638

Ion Ratio Lower Upper

42 100

74 115.7 91.7 137.5

44 7.6 5.6 8.4









#5
2-Fluorophenol
Concen: 41.752 ng
RT: 5.581 min Scan# 51
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:112 Resp: 444546
Ion Ratio Lower Upper
112 100
64 66.4 53.1 79.7
63 33.1 26.6 40.0



#6
Aniline
Concen: 20.505 ng
RT: 6.622 min Scan# 753
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion: 93 Resp: 382991
Ion Ratio Lower Upper
93 100
66 41.0 33.2 49.8
65 20.7 16.9 25.3








#7
Phenol-d6
Concen: 41.668 ng
RT: 6.575 min Scan# 747
Delta R.T. -0.012 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion: 99 Resp: 558427
Ion Ratio Lower Upper
99 100
42 21.5 17.0 25.6
71 33.9 27.7 41.5



#8
2-Chlorophenol
Concen: 20.702 ng
RT: 6.745 min Scan# 774
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion: 128 Resp: 231044
Ion Ratio Lower Upper
128 100
130 32.5 12.3 52.3
64 54.0 35.9 75.9









#9
Benzaldehyde
Concen: 19.504 ng
RT: 6.510 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 196003
Ion Ratio Lower Upper
77 100
105 95.3 76.8 116.8
106 92.0 74.4 114.4



#10
Phenol
Concen: 22.068 ng
RT: 6.593 min Scan# 748
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion: 94 Resp: 322193
Ion Ratio Lower Upper
94 100
65 29.0 10.0 50.0
66 38.8 20.8 60.8





#11

bis(2-Chloroethyl)ether

Concen: 20.792 ng

RT: 6.692 min Scan# 765

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion: 93 Resp: 232425

Ion Ratio Lower Upper

93 100

63 80.9 63.4 103.4

95 31.9 12.4 52.4



#12

1,3-Dichlorobenzene

Concen: 20.763 ng

RT: 6.904 min Scan# 801

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion: 146 Resp: 251734

Ion Ratio Lower Upper

146 100

148 62.7 50.6 76.0

75 33.7 27.4 41.2








#13
1,4-Dichlorobenzene
Concen: 20.733 ng
RT: 6.981 min Scan# 814
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:146 Resp: 253137
Ion Ratio Lower Upper
146 100
148 62.9 50.3 75.5
111 43.0 34.4 51.6



#14
1,2-Dichlorobenzene
Concen: 20.963 ng
RT: 7.134 min Scan# 840
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:146 Resp: 243437
Ion Ratio Lower Upper
146 100
148 63.4 50.7 76.1
111 44.0 36.6 54.8









#15
Benzyl Alcohol
Concen: 20.524 ng
RT: 7.092 min Scan# 811
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion: 79 Resp: 210016
Ion Ratio Lower Upper
79 100
108 75.0 61.1 91.7
77 60.5 48.9 73.3



#16
2,2'-oxybis(1-chloropropane)
Concen: 20.680 ng
RT: 7.234 min Scan# 857
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion: 45 Resp: 428827
Ion Ratio Lower Upper
45 100
77 11.2 0.0 31.2
79 8.4 0.0 28.5








#17
2-Methylphenol
Concen: 20.479 ng
RT: 7.198 min Scan# 851
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:	107	Resp:	192172
Ion	Ratio	Lower	Upper
107	100		
108	114.3	91.8	137.6
77	45.6	37.0	55.6
79	45.8	36.9	55.3



#18
Hexachloroethane
Concen: 20.828 ng
RT: 7.481 min Scan# 899
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:	117	Resp:	88036
Ion	Ratio	Lower	Upper
117	100		
119	95.5	77.8	116.8
201	94.8	77.0	115.4









#19
n-Nitroso-di-n-propylamine
Concen: 20.137 ng
RT: 7.363 min Scan# 81
Delta R.T. -0.012 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:	70	Resp:	173142
Ion	Ratio	Lower	Upper
70	100		
42	68.6	52.5	78.7
101	9.4	7.6	11.4
130	19.7	16.1	24.1



#20
3+4-Methylphenols
Concen: 21.745 ng
RT: 7.351 min Scan# 877
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:	107	Resp:	247578
Ion	Ratio	Lower	Upper
107	100		
108	90.7	70.0	110.0
77	70.4	57.8	97.8
79	29.5	9.5	49.5








#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.245 min Scan# 1029
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:	136	Resp:	642580
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	13.0
54	10.2	8.2	12.2
68	6.5	5.4	8.2



#22
Acetophenone
Concen: 21.151 ng
RT: 7.363 min Scan# 879
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:	105	Resp:	321774
Ion	Ratio	Lower	Upper
105	100		
71	9.0	5.8	8.6
51	36.8	30.2	45.2
120	22.9	18.2	27.4





#23

Nitrobenzene-d5

Concen: 43.013 ng

RT: 7.516 min Scan# 905

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion: 82 Resp: 554482

Ion Ratio Lower Upper

82 100

128 41.9 33.6 50.4

54 59.8 47.1 70.7



#24

Nitrobenzene

Concen: 20.739 ng

RT: 7.534 min Scan# 908

Delta R.T. -0.012 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion: 77 Resp: 249251


Ion Ratio Lower Upper

77 100

123 43.5 36.1 54.1

65 14.1 11.4 17.2








#25
Isophorone
Concen: 20.088 ng
RT: 7.775 min Scan# 949
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDIC020

Tgt Ion: 82 Resp: 457146
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 17.7 13.8 20.8



#26
2-Nitrophenol
Concen: 20.261 ng
RT: 7.851 min Scan# 962
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:139 Resp: 105681
Ion Ratio Lower Upper
139 100
109 29.0 23.6 35.4
65 55.2 42.6 64.0









#27
2,4-Dimethylphenol
Concen: 20.594 ng
RT: 7.887 min Scan# 90
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:122 Resp: 218960
Ion Ratio Lower Upper
122 100
107 114.5 91.8 137.6
121 56.1 46.3 69.5



#28
bis(2-Chloroethoxy)methane
Concen: 20.811 ng
RT: 7.987 min Scan# 985
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion: 93 Resp: 283946
Ion Ratio Lower Upper
93 100
95 32.2 25.8 38.6
123 11.6 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 20.569 ng

RT: 8.092 min Scan# 1003

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020




Tgt Ion:162 Resp: 184432

Ion Ratio Lower Upper

162 100

164 64.4 44.2 84.2

98 38.8 17.9 57.9



#30

1,2,4-Trichlorobenzene

Concen: 21.070 ng

RT: 8.181 min Scan# 1018

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:180 Resp: 204103

Ion Ratio Lower Upper

180 100

182 93.3 76.2 114.2

145 32.2 25.0 37.6








#31
Naphthalene
Concen: 20.949 ng
RT: 8.263 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:128 Resp: 662968
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.4
127 13.1 10.6 16.0



#32
Benzoic acid
Concen: 18.667 ng
RT: 7.957 min Scan# 980
Delta R.T. -0.035 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:122 Resp: 90732
Ion Ratio Lower Upper
122 100
105 126.7 104.8 144.8
77 99.7 77.6 117.6









#33
4-Chloroaniline
Concen: 20.825 ng
RT: 8.304 min Scan# 1039
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt	Ion:127	Resp:	269101
Ion	Ratio	Lower	Upper
127	100		
129	32.2	26.0	39.0
65	34.8	27.4	41.2
92	21.1	16.8	25.2



#34
Hexachlorobutadiene
Concen: 20.722 ng
RT: 8.386 min Scan# 1053
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:225	Resp: 122624
Ion Ratio	Lower Upper
225 100	
223 64.3	49.8 74.6
227 64.1	50.2 75.4









#35
 Caprolactam
 Concen: 19.911 ng
 RT: 8.651 min Scan# 1109
 Delta R.T. -0.029 min
 Lab File: BF143143.D
 Acq: 17 Jul 2025 12:34




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020


Tgt Ion:113 Resp: 53950
 Ion Ratio Lower Upper
 113 100
 55 202.9 174.0 214.0
 56 152.2 126.6 166.6



#36
 4-Chloro-3-methylphenol
 Concen: 20.246 ng
 RT: 8.775 min Scan# 1119
 Delta R.T. -0.006 min
 Lab File: BF143143.D
 Acq: 17 Jul 2025 12:34

Tgt Ion:107 Resp: 197317
 Ion Ratio Lower Upper
 107 100
 144 24.6 19.8 29.6
 142 76.4 60.6 90.8





#37

2-Methylnaphthalene

Concen: 20.943 ng

RT: 8.957 min Scan# 1150

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :





SSTDICC020

Tgt Ion:142 Resp: 403298

Ion Ratio Lower Upper

142 100

141 87.5 69.9 104.9



#38

1-Methylnaphthalene

Concen: 21.060 ng

RT: 9.057 min Scan# 1167

Delta R.T. 0.000 min

Lab File: BF143143.D




Acq: 17 Jul 2025 12:34


Tgt Ion:142 Resp: 417629

Ion Ratio Lower Upper

142 100

141 90.0 72.3 108.5





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.998 min Scan# 1178

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:164 Resp: 329670

Ion Ratio Lower Upper

164 100

162 97.5 79.0 118.6

160 43.2 35.1 52.7



#40

1,2,4,5-Tetrachlorobenzene

Concen: 20.804 ng

RT: 9.122 min Scan# 1178

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Tgt Ion:216 Resp: 198009




Ion Ratio Lower Upper

216 100

214 79.4 63.8 95.8

179 21.0 16.4 24.6

108 19.9 15.6 23.4









#41
Hexachlorocyclopentadiene
Concen: 19.848 ng
RT: 9.116 min Scan# 1177
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:237 Resp: 122922
Ion Ratio Lower Upper
237 100
235 63.2 43.0 83.0
272 13.2 0.0 33.3



#42
2,4,6-Tribromophenol
Concen: 41.034 ng
RT: 10.780 min Scan# 1460
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:330 Resp: 139748
Ion Ratio Lower Upper
330 100
332 97.9 76.7 115.1
141 30.4 23.3 34.9









#43
2,4,6-Trichlorophenol
Concen: 20.601 ng
RT: 9.228 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:196 Resp: 135703
Ion Ratio Lower Upper
196 100
198 95.4 76.2 114.2
200 29.9 23.9 35.9



#44
2,4,5-Trichlorophenol
Concen: 20.584 ng
RT: 9.263 min Scan# 1202
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:196 Resp: 135634
Ion Ratio Lower Upper
196 100
198 95.3 77.3 115.9
97 51.2 40.3 60.5
132 29.3 23.1 34.7





#45

2-Fluorobiphenyl

Concen: 42.301 ng

RT: 9.316 min Scan# 1211

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:172 Resp: 1019688

Ion Ratio Lower Upper

172 100

171 35.2 28.6 42.8

170 23.2 18.8 28.2



#46

1,1'-Biphenyl

Concen: 21.019 ng

RT: 9.416 min Scan# 1228

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:154 Resp: 540637

Ion Ratio Lower Upper

154 100

153 40.2 20.3 60.3

76 13.9 0.0 33.3









#47
2-Chloronaphthalene
Concen: 20.650 ng
RT: 9.445 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:162 Resp: 393002
Ion Ratio Lower Upper
162 100
127 40.6 33.4 50.2
164 32.9 26.2 39.2



#48
2-Nitroaniline
Concen: 20.528 ng
RT: 9.528 min Scan# 1247
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion: 65 Resp: 122517
Ion Ratio Lower Upper
65 100
92 65.6 52.4 78.6
138 100.8 82.3 123.5





#49

Acenaphthylene

Concen: 20.789 ng

RT: 9.857 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:152 Resp: 663048

Ion Ratio Lower Upper

152 100

151 20.2 16.2 24.2

153 13.3 10.6 16.0



#50

Dimethylphthalate

Concen: 20.558 ng

RT: 9.710 min Scan# 1278

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:163 Resp: 441768


Ion Ratio Lower Upper

163 100

194 4.4 3.4 5.2

164 10.3 8.3 12.5









#51
2,6-Dinitrotoluene
Concen: 20.529 ng
RT: 9.769 min Scan# 1289
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:165 Resp: 89742
Ion Ratio Lower Upper
165 100
63 63.2 49.4 74.2
89 60.6 46.2 69.2



#52
Acenaphthene
Concen: 20.702 ng
RT: 10.033 min Scan# 1333
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:154 Resp: 390249
Ion Ratio Lower Upper
154 100
153 107.3 86.6 129.8
152 50.9 40.4 60.6








#53
3-Nitroaniline
Concen: 20.314 ng
RT: 9.933 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:138 Resp: 103870
Ion Ratio Lower Upper
138 100
108 11.7 9.0 13.6
92 123.4 93.0 139.4



#54
2,4-Dinitrophenol
Concen: 19.073 ng
RT: 10.039 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:184 Resp: 26820
Ion Ratio Lower Upper
184 100
63 145.3 81.8 122.6#
154 980.0 412.2 618.4#









#55
Dibenzofuran
Concen: 20.806 ng
RT: 10.204 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:168 Resp: 582320
Ion Ratio Lower Upper
168 100
139 35.9 29.5 44.3
169 13.1 10.9 16.3



#56
4-Nitrophenol
Concen: 19.715 ng
RT: 10.081 min Scan# 1341
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:139 Resp: 75275
Ion Ratio Lower Upper
139 100
109 68.3 48.5 88.5
65 105.1 83.4 123.4









#57
2,4-Dinitrotoluene
Concen: 20.577 ng
RT: 10.169 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 112851
Ion Ratio Lower Upper
165 100
63 49.9 41.0 61.4
89 77.0 61.1 91.7
182 2.9 2.3 3.5



#58
Fluorene
Concen: 20.750 ng
RT: 10.545 min Scan# 1420
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:166 Resp: 435866
Ion Ratio Lower Upper
166 100
165 92.1 73.7 110.5
167 13.8 10.6 16.0









#59
2,3,4,6-Tetrachlorophenol
Concen: 20.859 ng
RT: 10.316 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:232 Resp: 113851
Ion Ratio Lower Upper
232 100
131 44.3 35.3 52.9
130 2.1 1.7 2.5
166 29.1 23.4 35.2



#60
Diethylphthalate
Concen: 20.719 ng
RT: 10.410 min Scan# 1397
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:149 Resp: 440058
Ion Ratio Lower Upper
149 100
177 21.7 17.5 26.3
150 12.5 10.2 15.2









#61
4-Chlorophenyl-phenylether
Concen: 20.699 ng
RT: 10.539 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:204 Resp: 216496
Ion Ratio Lower Upper
204 100
206 32.4 25.8 38.8
141 57.8 47.0 70.4



#62
4-Nitroaniline
Concen: 20.005 ng
RT: 10.545 min Scan# 1420
Delta R.T. -0.012 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:138 Resp: 87666
Ion Ratio Lower Upper
138 100
92 50.7 29.3 69.3
108 69.0 48.3 88.3





#63

Azobenzene

Concen: 20.690 ng

RT: 10.698 min Scan# 1446

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion: 77 Resp: 458013





Ion Ratio Lower Upper

77 100

182 25.8 4.6 44.6

105 21.8 1.4 41.4

51 33.5 13.7 53.7



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.480 min Scan# 1579

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:188 Resp: 532921


Ion Ratio Lower Upper

188 100

94 11.0 8.3 12.5

80 11.9 9.0 13.6









#65
4,6-Dinitro-2-methylphenol
Concen: 20.060 ng
RT: 10.580 min Scan# 1427
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:198 Resp: 46122
Ion Ratio Lower Upper
198 100
51 52.7 37.2 77.2
105 43.1 24.8 64.8



#66
n-Nitrosodiphenylamine
Concen: 20.488 ng
RT: 10.651 min Scan# 1438
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:169 Resp: 384467
Ion Ratio Lower Upper
169 100
168 65.6 52.1 78.1
167 35.4 28.2 42.2





#67

4-Bromophenyl-phenylether

Concen: 20.259 ng

RT: 11.028 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:248 Resp: 128663

Ion Ratio Lower Upper

248 100

250 97.8 78.3 117.5

141 72.3 59.2 88.8



#68

Hexachlorobenzene

Concen: 20.191 ng

RT: 11.098 min Scan# 1514

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:284 Resp: 134031


Ion Ratio Lower Upper

284 100

142 32.7 27.0 40.6

249 31.5 25.0 37.6





#69

Atrazine

Concen: 20.370 ng

RT: 11.169 min Scan# 1526

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020




Tgt Ion:200 Resp: 105386

Ion Ratio Lower Upper

200 100

173 27.7 7.9 47.9

215 51.0 30.1 70.1



#70

Pentachlorophenol

Concen: 19.957 ng

RT: 11.280 min Scan# 1545

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:266 Resp: 77906


Ion Ratio Lower Upper

266 100

268 64.3 51.9 77.9

264 62.6 51.0 76.6









#71
Phenanthrene
Concen: 21.042 ng
RT: 11.510 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:178 Resp: 595411
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.2 12.5 18.7



#72
Anthracene
Concen: 20.944 ng
RT: 11.557 min Scan# 1592
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34


Tgt Ion:178 Resp: 604416
Ion Ratio Lower Upper
178 100
176 18.0 15.0 22.4
179 15.4 12.5 18.7








#73
Carbazole
Concen: 20.769 ng
RT: 11.704 min Scan# 1617
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC020







Tgt Ion:167 Resp: 523365
Ion Ratio Lower Upper
167 100
166 21.2 17.0 25.6
139 12.7 10.0 15.0



#74
Di-n-butylphthalate
Concen: 21.108 ng
RT: 12.039 min Scan# 1674
Delta R.T. 0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:149 Resp: 601867
Ion Ratio Lower Upper
149 100
150 9.2 7.4 11.2
104 4.7 3.8 5.8





#75

Fluoranthene

Concen: 20.894 ng

RT: 12.692 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:202 Resp: 542662

Ion Ratio Lower Upper

202 100

101 12.5 0.0 31.7

203 17.7 0.0 37.4



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.121 min Scan# 2028

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:240 Resp: 305398


Ion Ratio Lower Upper

240 100

120 11.3 9.4 14.2

236 25.5 20.2 30.4









#77
Benzidine
Concen: 21.919 ng
RT: 12.810 min Scan# 1805
Delta R.T. -0.000 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:184 Resp: 257904
Ion Ratio Lower Upper
184 100
185 14.9 12.2 18.4
183 10.7 9.0 13.4



#78
Pyrene
Concen: 20.897 ng
RT: 12.922 min Scan# 1824
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:202 Resp: 544650
Ion Ratio Lower Upper
202 100
200 19.1 15.7 23.5
203 17.6 14.2 21.2





#79

Terphenyl-d14

Concen: 42.238 ng

RT: 13.063 min Scan# 1848

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:244 Resp: 866830

Ion Ratio Lower Upper

244 100

212 6.3 5.2 7.8

122 11.3 8.2 12.2



#80

Butylbenzylphthalate

Concen: 20.491 ng

RT: 13.539 min Scan# 1929

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:149 Resp: 163546


Ion Ratio Lower Upper

149 100

91 71.2 57.9 86.9

206 21.4 17.0 25.4





#81

Benzo(a)anthracene

Concen: 20.629 ng

RT: 14.110 min Scan# 2026

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:228 Resp: 421791

Ion Ratio Lower Upper

228 100

226 26.6 21.7 32.5

229 19.8 15.8 23.6



#82

3,3'-Dichlorobenzidine

Concen: 22.181 ng

RT: 14.068 min Scan# 2019

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:252 Resp: 151153


Ion Ratio Lower Upper

252 100

254 64.9 51.7 77.5

126 12.0 9.4 14.0





#83

Chrysene

Concen: 20.714 ng

RT: 14.145 min Scan# 2033

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:228 Resp: 380799

Ion Ratio Lower Upper

228 100

226 29.3 23.8 35.8

229 19.9 15.6 23.4



#84

Bis(2-ethylhexyl)phthalate

Concen: 20.839 ng

RT: 14.098 min Scan# 2024

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:149 Resp: 251739

Ion Ratio Lower Upper

149 100

167 28.2 22.3 33.5

279 4.3 3.6 5.4





#85

Di-n-octyl phthalate

Concen: 19.916 ng

RT: 14.716 min Scan# 2129

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDIC020





Tgt Ion:149 Resp: 436102

Ion Ratio Lower Upper

149 100

167 1.5 1.1 1.7

43 12.4 9.7 14.5



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.621 min Scan# 2283

Delta R.T. 0.000 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:264 Resp: 338518


Ion Ratio Lower Upper

264 100

260 23.1 18.5 27.7

265 21.7 17.4 26.2









#87
Indeno(1,2,3-cd)pyrene
Concen: 20.286 ng
RT: 17.151 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34




Instrument :
BNA_F
ClientSampleId :
SSTDICC020


Tgt Ion:276 Resp: 484240
Ion Ratio Lower Upper
276 100
138 27.8 21.8 32.6
277 25.3 20.0 30.0



#88
Benzo(b)fluoranthene
Concen: 21.071 ng
RT: 15.174 min Scan# 2207
Delta R.T. -0.006 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:252 Resp: 435747
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 10.4 8.0 12.0





#89

Benzo(k)fluoranthene

Concen: 20.044 ng

RT: 15.204 min Scan# 2212

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34

Instrument :

BNA_F

ClientSampleId :

SSTDICC020





Tgt Ion:252 Resp: 369899

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.6

125 10.2 8.1 12.1



#90

Benzo(a)pyrene

Concen: 20.371 ng

RT: 15.557 min Scan# 2272

Delta R.T. -0.006 min

Lab File: BF143143.D

Acq: 17 Jul 2025 12:34




Tgt Ion:252 Resp: 388136


Ion Ratio Lower Upper

252 100

253 21.2 17.3 25.9

125 11.0 8.6 13.0









#91
Dibenzo(a,h)anthracene
Concen: 20.534 ng
RT: 17.168 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:278 Resp: 398949
Ion Ratio Lower Upper
278 100
139 18.2 14.1 21.1
279 23.4 18.6 28.0



#92
Benzo(g,h,i)perylene
Concen: 20.534 ng
RT: 17.615 min Scan# 2622
Delta R.T. -0.012 min
Lab File: BF143143.D
Acq: 17 Jul 2025 12:34

Tgt Ion:276 Resp: 385915
Ion Ratio Lower Upper
276 100
277 23.3 19.1 28.7
138 23.3 18.9 28.3

