

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: <u>Q2646</u>	Analytical Method: <u>8270E</u>
Client: <u>Environmental Restoration, LLC</u>	DataFile: <u>BF143213.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168954BS	Pyridine	50	37.0	ug/L	74				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	42.5	ug/L	85				70 (76)	130 (103)	
	2-Methylphenol	50	43.9	ug/L	88				70 (69)	130 (109)	
	3+4-Methylphenols	50	42.7	ug/L	85				20 (67)	160 (106)	
	Hexachloroethane	50	44.3	ug/L	89				20 (76)	160 (118)	
	Nitrobenzene	50	44.2	ug/L	88				70 (58)	130 (106)	
	Hexachlorobutadiene	50	44.2	ug/L	88				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	44.7	ug/L	89				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	44.7	ug/L	89				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	53.0	ug/L	106				70 (60)	130 (115)	
	Hexachlorobenzene	50	43.9	ug/L	88				70 (73)	130 (106)	
	Pentachlorophenol	100	61.8	ug/L	62				20 (47)	160 (114)	