

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143188.D
Instrument ID: BNA_F

Contract: ENVI60
SDG NO.: Q2646
DFTPP Injection Date: 07/22/2025
DFTPP Injection Time: 14:54

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	77.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143189.D	07/22/2025	15:23
PB168954BL	PB168954BL	BF143190.D	07/22/2025	15:52
FRAC TANK	Q2646-03	BF143198.D	07/22/2025	19:51
FRAC TANKMS	Q2646-03MS	BF143199.D	07/22/2025	20:20
FRAC TANKMSD	Q2646-03MSD	BF143200.D	07/22/2025	20:49