

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2646</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Environmental Restoration, LLC</u>	Datafile :	<u>VN087371.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0721WBS01	Vinyl chloride	20	19.0	ug/L	95			70 (65)	130 (117)	
	1,1-Dichloroethene	20	18.1	ug/L	91			70 (74)	130 (110)	
	2-Butanone	100	86.4	ug/L	86			40 (65)	160 (122)	
	Carbon Tetrachloride	20	19.6	ug/L	98			70 (77)	130 (113)	
	Chloroform	20	17.9	ug/L	90			70 (79)	130 (113)	
	Benzene	20	19.4	ug/L	97			70 (82)	130 (109)	
	1,2-Dichloroethane	20	18.3	ug/L	92			70 (80)	130 (115)	
	Trichloroethene	20	18.8	ug/L	94			70 (77)	130 (113)	
	Tetrachloroethene	20	18.3	ug/L	92			70 (67)	130 (123)	
	Chlorobenzene	20	18.6	ug/L	93			70 (82)	130 (109)	